

<b>1.01.20</b>	<b>Continuous Glucose Monitoring</b>		
<b>Original Policy Date:</b>	March 1, 2020	<b>Effective Date:</b>	June 1, 2022
<b>Section:</b>	1.0 Durable Medical Equipment	<b>Page:</b>	Page 1 of 65

## Policy Statement

[Continuous glucose monitoring](#) (CGM) of glucose levels in interstitial fluid, as a technique of diabetic monitoring, may be considered **medically necessary** when **both** of the following situations occur:

- I. Patients with insulin dependent (type 1 or type 2) diabetes requires multiple (three or more) daily doses of insulin
- II. The device includes an audible or tactile (vibrating) alarm for low glucose alerts without patient intervention  
(NOTE: the FreeStyle Libre 14 day device does not have alarms but the FreeStyle Libre 2 does have appropriate alarms, as do Dexcom G5 and G6)

The use of [implantable CGM](#) devices (e.g., Eversense®) are considered **investigational** (see Policy Guidelines section).

The use of [continuous noninvasive glucose monitoring devices](#) (see Policy Guidelines section) are considered **investigational**.

The use of stand-alone CGMs (e.g., Guardian Connect, that is not part of an insulin infusion pump) [adjunctive or non-therapeutic devices](#) are considered **investigational**.

Other uses of long-term CGM of glucose levels in interstitial fluid as a technique of diabetic monitoring in patients who are not insulin dependent (including use in gestational diabetes) are considered **investigational**.

**NOTE:** Refer to [Appendix A](#) to see the policy statement changes (if any) from the previous version.

## Policy Guidelines

This policy only evaluates continuous (real time or intermittent) interstitial glucose monitors and does not evaluate the use of continuous glucose monitoring that is integrated into insulin pumps.

Supplies needed for continuous glucose monitoring such as disposable sensors or transmitters are not covered for patients who do not meet the criteria for CGM.

### Continuous Glucose Monitors

A separate monitor device was needed in the past but now there are options for phone apps that can be used instead of the monitor. Approval would include either the app or the monitor (receiver) device.

### Adjunctive/non-therapeutic and non-adjunctive/therapeutic

Devices that do not require separate fingerstick glucose checks to make treatment decisions are considered to be “therapeutic” or “non-adjunctive”. Those that do require separate fingerstick glucose checks to confirm calibration and before making treatment decisions are “adjunctive” or “non-therapeutic” (meaning they are adjuncts or add-ons to fingersticks but not independent of them).

- Adjunctive examples: Guardian Connect by Medtronic
- Non-adjunctive examples: Dexcom G5 and G6; Freestyle Libre 14 day and 2; Eversense

**Real time** (rtCGM) devices automatically record and store data continuously with no manual intervention needed. Examples include but are not limited to Dexcom G6 and Guardian Connect, both of which also include alarm systems.

**Intermittently scanned or flash devices**

Intermittently scanned devices (flash CGM or isCGM) require manually bringing the monitor physically close to the sensor to record and display current data (thus, intermittent based on how often the levels are checked/recorded). Examples include but are not limited to the Freestyle Libre 2 which has both alarms and the ability to store data for 8 hours; but data can be lost if not recorded for more than 8 hours. The Freestyle Libre 14 day (original version) does not have alarm capability.

**Implantable CGM** devices (e.g., Eversense) are small and inserted under the skin through a small incision, then removed and replaced with a new device in another location every 90 days (a new XL version is implanted every 180 days instead).

Continuous noninvasive devices are often worn like a watch, with a skin patch or with a fingerprint type device and have special sensors that do not require the tiny subcutaneous catheters used with traditional CGMs. Without breaking the skin, this sensor measures the patient's blood sugar levels.

Disposable durable medical equipment (DME) supplies such as sensors and transmitters are considered a covered benefit when the CGM device or app has also been approved. Blue Shield of California (BSC) plans exclude coverage of supplies for units purchased by the patient without prior authorization. Please check benefit plan descriptions for details.

Best practices in diabetes control include compliance with a self-monitoring blood glucose regimen of four or more fingersticks each day and use of an insulin pump or multiple daily injections of insulin. During pregnancy, three or more insulin injections daily could also be considered best practice for patients not on an insulin pump prior to the pregnancy. Prior short-term (72-hour) use of an intermittent glucose monitor would be considered a part of best practices for those considering long-term use of a continuous glucose monitor. Women with type 1 diabetes taking insulin who are pregnant or about to become pregnant with poorly controlled diabetes are another subset of patients that might require either short or long term continuous glucose monitoring.

Significant hypoglycemia may include recurrent, unexplained, severe (generally blood glucose levels less than 50 mg/dL) hypoglycemia or impaired awareness of hypoglycemia that puts the patient or others at risk. When this happens at night while the patient is asleep, the alarm functions are even more important.

The strongest evidence exists for use of continuous glucose monitoring (CGM) devices in patients age 25 and older. However, age may be a proxy for motivation and good control of disease, so it is also reasonable to select patients based on their ability to self-manage their disease, rather than their age. Most CGM devices have U.S. Food and Drug Administration labeling related to age.

Providers board-certified in endocrinology and/or providers with a focus on the practice of diabetes care may be considered qualified to evaluate and oversee individuals for continuous (i.e., long-term) monitoring.

**Coding**

The language of the CPT codes that specifically described monitoring of glucose levels in the interstitial fluid using implanted devices was revised to state that the devices are used for a minimum of 72 hours:

- **95250:** Ambulatory continuous glucose monitoring of interstitial tissue fluid via a subcutaneous sensor for a minimum of 72 hours; physician or other qualified health care professional (office) provided equipment, sensor placement, hook-up, calibration of monitor, patient training, removal of sensor, and printout of recording
- **95251:** Ambulatory continuous glucose monitoring of interstitial tissue fluid via a subcutaneous sensor for a minimum of 72 hours; analysis, interpretation and report

The following CPT code specifically describes glucose monitoring on equipment provided by the patient:

- **95249:** Ambulatory continuous glucose monitoring of interstitial tissue fluid via a subcutaneous sensor for a minimum of 72 hours; patient-provided equipment, sensor placement, hook-up, calibration of monitor, patient training, and printout of recording

CPT code 99091 might also be used for this monitoring:

- **99091:** Collection and interpretation of physiologic data (e.g., ECG, blood pressure, glucose monitoring) digitally stored and/or transmitted by the patient and/or caregiver to the physician or other qualified health care professional, qualified by education, training, licensure/regulation (when applicable) requiring a minimum of 30 minutes of time, each 30 days

*Effective April 1, 2022*, the Centers for Medicare & Medicaid created 2 HCPCS codes specific to the use of devices that cannot be used to make treatment decisions (adjunctive or non-therapeutic) (see Continuous Glucose Monitors section above in Guidelines):

- **A4238:** Supply allowance for adjunctive continuous glucose monitor (CGM), includes all supplies and accessories, 1 month supply = 1 unit of service
- **E2102:** Adjunctive continuous glucose monitor or receiver

The Centers for Medicare & Medicaid created 2 HCPCS codes specific to the use of devices to make treatment decisions (non-adjunctive or therapeutic (see Continuous Glucose Monitors section above in Guidelines):

- **K0553:** Supply allowance for therapeutic continuous glucose monitor (CGM), includes all supplies and accessories, 1 month supply = 1 unit of service
- **K0554:** Receiver (monitor), dedicated, for use with therapeutic glucose continuous monitor system

The following non-specific HCPCS codes have been deleted by CMS but are still available for use otherwise for continuous glucose monitoring systems. These are non-specific related to adjunctive and non-adjunctive, so CMS prefers the use of the K codes (K0553 and K0554) as well as the new codes (A4238 and E2102) that are more specific:

- **A9276:** Sensor; invasive (e.g., subcutaneous), disposable, for use with interstitial continuous glucose monitoring system, 1 unit = 1 day supply
- **A9277:** Transmitter; external, for use with interstitial continuous glucose monitoring system

## Description

Tight glucose control in patients with diabetes has been associated with improved health outcomes. Several devices are available to measure glucose levels automatically and frequently (e.g., every 5 to 10 minutes). The devices measure glucose in the interstitial fluid and are approved as adjuncts to or replacements for traditional self-monitoring of blood glucose levels. Devices can be used on a long-term (continuous) or short-term (often referred to as intermittent) basis.

## Related Policies

- N/A

## Benefit Application

Benefit determinations should be based in all cases on the applicable contract language. To the extent there are any conflicts between these guidelines and the contract language, the contract language will control. Please refer to the member's contract benefits in effect at the time of service to determine coverage or non-coverage of these services as it applies to an individual member.

Some state or federal mandates (e.g., Federal Employee Program [FEP]) prohibits plans from denying Food and Drug Administration (FDA)-approved technologies as investigational. In these instances, plans may have to consider the coverage eligibility of FDA-approved technologies on the basis of medical necessity alone.

## Regulatory Status

Multiple CGM systems have been approved by the FDA through the premarket approval process (see Table 1). FDA product codes: QCD, MDS.

CGM devices labeled as "Pro" for specific professional use with customized software and transmission to health care professionals are not enumerated in this list. The Flash glucose monitors (eg FreeStyle Libre, Abbott) use intermittent scanning and do not have continuous or real-time alerts.

**Table 1. CGM Systems Approved by the U.S. Food and Drug Administration**

Device	Manufacturer	Approval	Indications
Continuous Glucose Monitoring System (CGMS®)	MiniMed	1999	3-d use in physician's office
GlucoWatch G2® Biographer		2001	Not available since 2008
Guardian®-RT (Real-Time) CGMS	MiniMed (now Medtronic)	2005	
Dexcom® STS CGMS system	Dexcom	2006	
Paradigm® REAL-Time System (second-generation called Paradigm Revel System)	MiniMed (now Medtronic)	2006	Integrates CGM with a Paradigm insulin pump
FreeStyle Navigator® CGM System	Abbott	2008	
Dexcom® G4 Platinum	Dexcom	2012	Adults ≥18 y; can be worn for up to 7 d
Dexcom® G5 Mobile CGM	Dexcom	2014	Expanded to include patients with diabetes 2-17 y
Dexcom® G6 Continuous Glucose Monitoring System	Dexcom	2018	Replacement for fingerstick blood glucose testing in patients ≥2 y. System requires at least 2 daily fingerstick tests for calibration purposes, but additional fingersticks are not necessary because treatment decisions can be made based on device readings <sup>5</sup> . Indicated for the management of diabetes in persons age ≥2 years. Intended to replace fingerstick blood glucose testing for diabetes treatment decisions. Intended to autonomously communicate with

Device	Manufacturer	Approval	Indications
Freestyle Libre® Flash Glucose Monitoring System	Abbott	2017	digitally connected devices, including automated insulin dosing (AID) systems. with 10-day wear Adults ≥18 y. Indicated for the management of diabetes and can be worn up to 10 days It is designed to replace blood glucose testing for diabetes treatment decisions.
Freestyle Libre® Flash Glucose Monitoring System	Abbott	2018	Adults ≥18 y. Extended duration of use to 14 days
Guardian Connect	Medtronic MiniMed	2018	Adolescents and adults (14-75 years) Continuous or periodic monitoring of interstitial glucose levels. Provides real-time glucose values, trends, and alerts through a Guardian Connect app installed on a compatible consumer electronic mobile device Adults ≥18 y.
Eversense Continuous Glucose Monitoring System	Senseonics	2018 2019	Continually measuring glucose levels up to 90 days. Use as an adjunctive device to complement, not replace, information obtained from standard home blood glucose monitoring devices. Adults ≥18 y. Continually measuring glucose levels up to 90 days. Indicated for use to replace fingerstick blood glucose measurements for diabetes treatment decisions. Historical data from the system can be interpreted to aid in providing therapy adjustments.

CGM: continuous glucose monitoring.

<sup>a</sup> As a supplement to the G4 premarketing approval.

Food and Drug Administration product codes: MDS, PQF, QCD.

## Rationale

### Background Blood Glucose Control

The advent of blood glucose monitors for use by patients in the home revolutionized the management of diabetes. Using fingersticks, patients can monitor their blood glucose levels both to determine the adequacy of hyperglycemia control and to evaluate hypoglycemic episodes. Tight glucose control, defined as a strategy involving frequent glucose checks and a target hemoglobin A<sub>1c</sub> (HbA<sub>1c</sub>) level in the range of 7%, is now considered the standard of care for diabetic patients. Randomized controlled trials assessing tight control have demonstrated benefits for patients with type 1 diabetes in decreasing microvascular complications. The impact of tight control on type 1 diabetes and macrovascular complications such as stroke or myocardial infarction is less certain. The Diabetes Control and Complications Trial (2002) demonstrated that a relative HbA<sub>1c</sub> level reduction of 10% is clinically meaningful and corresponds to approximately a 40% decrease in risk for progression of diabetic retinopathy and a 25% decrease in risk for progression of renal disease.<sup>1</sup>

Due to an increase in turnover of red blood cells during pregnancy, HbA<sub>1c</sub> levels are slightly lower in women with a normal pregnancy compared with nonpregnant women. The target A<sub>1c</sub> in women with diabetes is also lower in pregnancy. The American Diabetes Association recommends that, if achievable without significant hypoglycemia, the A<sub>1c</sub> levels should range between 6.0% to 6.5%; an A<sub>1c</sub> level less than 6% may be optimal as the pregnancy progresses.<sup>2</sup>

Tight glucose control requires multiple daily measurements of blood glucose (i.e., before meals and at bedtime), a commitment that some patients may find difficult to meet. The goal of tight glucose control has to be balanced with an associated risk of hypoglycemia. Hypoglycemia is known to be a risk in patients with type 1 diabetes. While patients with insulin-treated type 2 diabetes may also experience severe hypoglycemic episodes, there is a lower relative likelihood

of severe hypoglycemia compared with patients who had type 1 diabetes.<sup>3,4</sup> An additional limitation of periodic self-measurements of blood glucose is that glucose levels are seen in isolation, and trends in glucose levels are undetected. For example, while a diabetic patient's fasting blood glucose level might be within normal values, hyperglycemia might be undetected postprandially, leading to elevated HbA1c levels.

### **Management**

Measurements of glucose in the interstitial fluid have been developed as a technique to measure glucose values automatically throughout the day, producing data that show the trends in glucose levels. Although devices measure glucose in the interstitial fluid on a periodic rather than a continuous basis, this type of monitoring is referred to as continuous glucose monitoring (CGM).

Currently, CGM devices are of 2 designs; real-time CGM (rtCGM) provides real-time data on glucose level, glucose trends, direction, and rate of change and, intermittently viewed (iCGM) devices that show continuous glucose measurements retrospectively. These devices are also known as flash-glucose monitors.

Approved devices now include devices indicated for pediatric use and those with more advanced software, more frequent measurements of glucose levels, or more sophisticated alarm systems. Devices initially measured interstitial glucose every 5 to 10 minutes and stored data for download and retrospective evaluation by a clinician. With currently available devices, the intervals at which interstitial glucose is measured range from every 1 to 2 minutes to 5 minutes, and most provide measurements in real-time directly to patients. While CGM potentially eliminates or decreases the number of required daily fingersticks, it should be noted that, according to the U.S. Food and Drug Administration (FDA) labeling, some marketed monitors are not intended as an alternative to traditional self-monitoring of blood glucose levels but rather as adjuncts to monitoring, supplying additional information on glucose trends not available from self-monitoring. The devices must be calibrated twice daily with blood glucose measurements from fingersticks, and are less reliable when used after exercise or post-prandial. Devices may be used intermittently (i.e., for periods of 72 hours) or continuously (i.e., on a long-term basis).

### **Literature Review**

Evidence reviews assess the clinical evidence to determine whether the use of technology improves the net health outcome. Broadly defined, health outcomes are the length of life, quality of life (QOL), and ability to function, including benefits and harms. Every clinical condition has specific outcomes that are important to patients and managing the course of that condition. Validated outcome measures are necessary to ascertain whether a condition improves or worsens; and whether the magnitude of that change is clinically significant. The net health outcome is a balance of benefits and harms.

To assess whether the evidence is sufficient to draw conclusions about the net health outcome of technology, 2 domains are examined: the relevance, and quality and credibility. To be relevant, studies must represent one or more intended clinical use of the technology in the intended population and compare an effective and appropriate alternative at a comparable intensity. For some conditions, the alternative will be supportive care or surveillance. The quality and credibility of the evidence depend on study design and conduct, minimizing bias and confounding that can generate incorrect findings. The randomized controlled trial (RCT) is preferred to assess efficacy; however, in some circumstances, nonrandomized studies may be adequate. RCTs are rarely large enough or long enough to capture less common adverse events and long-term effects. Other types of studies can be used for these purposes and to assess generalizability to broader clinical populations and settings of clinical practice.

Technical reliability is outside the scope of these reviews, and credible information on technical reliability is available from other sources.

The evidence review focuses on the clinical utility of continuous glucose monitoring (CGM) systems. That is, their ability to provide additional information on glucose levels leads to improved glucose control, or to reduce the morbidity and mortality associated with clinically significant severe and acute hypoglycemic or hyperglycemic events. Because diabetic control encompasses numerous variables, including the diabetic regimen and patient self-management, RCTs are important to isolate the contribution of interstitial glucose measurements to overall diabetes management.

For the evaluation of the clinical utility of CGM, studies would need to use the test as either an adjunct or a replacement to current disease status measures to manage treatment decisions in patients with diabetes. Outcomes would include measures of glucose control, QOL and measures of disease progression. Hemoglobin A1c (HbA1c) has commonly been accepted as a marker of glucose control; more recent studies have also reported time in hyperglycemia, time in hypoglycemia, and time in range as intermediate outcome measures.

### **Continuous Glucose Monitoring Devices for Long-Term Use in Type 1 Diabetes**

In some parts of the analysis of type 1 diabetes, BCBSA combines discussion of long-term and short-term glucose monitoring because several systematic reviews and RCTs provided information relevant to both indications.

#### **Clinical Context and Therapy Purpose**

The purpose of long-term CGM devices is to provide a testing option that is an alternative to or an improvement on existing testing used in the management of individuals with type 1 diabetes.

The question addressed in this evidence review is: Does long-term use of a CGM device improve the net health outcome for individuals with type 1 diabetes?

The following PICO was used to select literature to inform this review.

#### ***Populations***

The relevant population of interest is individuals with type 1 diabetes. All individuals with type 1 diabetes require engagement in a comprehensive self-management and clinical assessment program that includes assessment of blood glucose control.

#### ***Interventions***

The test being considered is the use of a CGM device to assess blood glucose levels as part of optimal diabetes management.

#### ***Comparators***

The following practice is currently being used to measure glucose levels: capillary blood sampling (finger stick) for self-monitoring of blood glucose (SMBG). Standard treatment for patients with type 1 diabetes includes injection of long-acting basal insulin plus multiple daily injections (MDI) of rapid-acting insulin boluses as required for meal intake. Activity level may require patients need to modify the timing and dose of insulin administration. Individuals with type 1 diabetes may also use an insulin pump either for initial treatment or convert to pump use after a period of MDI. Individuals are required to check their blood glucose before making preprandial insulin calculations, in response to symptoms of hypoglycemia or related to activity-related insulin adjustments.

#### ***Outcomes***

The general outcomes of interest are change in HbA1c levels, time spent in hypoglycemia and hyperglycemia, time in range (generally glucose of 70 to 180 mg/dl), the incidence of hypoglycemic events, complications of hypoglycemia, and QOL. To assess short-term outcomes such as HbA1c levels, a minimum follow-up of 8 to 12 weeks is appropriate.

## Study Selection

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs;
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess long-term outcomes and adverse effects, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

## Review of Evidence

### Systematic Reviews

A number of systematic reviews and meta-analyses have assessed RCTs evaluating CGM for long-term, daily use in treating type 1 diabetes.<sup>6,7,8,9,10,11</sup> These systematic reviews have focused on slightly different populations, and some did not separate long-term CGM from intermittent glucose monitoring.<sup>2</sup>

The only analysis to use individual patient data was published by Benkhadra et al (2017).<sup>12</sup> The meta-analysis evaluated data from 11 RCTs that enrolled patients with type 1 diabetes and compared real-time CGM with a control intervention. Studies in which patients used insulin pumps or received multiple daily insulin injections were included. Reviewers contacted corresponding study authors requesting individual patient data; data were not obtained for 1 trial. Mean baseline HbA1c levels were 8.2% in adults and 8.3% in children and adolescents. The overall risk of bias in the studies was judged to be moderate. In pooled analyses, there was a statistically significantly greater decrease in HbA1c levels with real-time CGM versus control conditions. Overall, the degree of difference between groups was 0.26%. In subgroup analyses by age, there was a significantly greater change in HbA1c levels among individuals 15 years and older, but not among the younger age groups. There were no significant differences between groups in the time spent in hypoglycemia or the incidence of hypoglycemic events. Key findings are shown in Table 2.

**Table 2. Individual Patient Data Meta-Analytic Outcomes for Real-Time CGM in Type 1 Diabetes**

No. of Trials	N	Group	Point Estimate	95% Confidence Intervals	p
<b>Change in HbA1c levels, %</b>					
8	1371	Overall	-0.258	0.464 to -0.052	.014
7	902	Age >15 y	-0.356	0.551 to -0.160	<.001
7	178	Age 13-15 y	-0.039	-0.320 to 0.242	.787
7	291	Age ≤12 y	-0.047	0.217 to 0.124	.592
<b>Time spent in hypoglycemia &lt;60 mg/dL, min</b>					
4	706	Overall	-8.549	-31.083 to 13.985	.457
4	467	Age >15 y	-8.095	-32.615 to 16.425	.518
3	109	Age 13-15 y	-13.966	31.782 to 3.852	.124
3	130	Age ≤12 y	-9.366	19.898 to 1.167	.081
<b>Incidence of hypoglycemic events &lt;70 mg/dL, mean no. events</b>					
3	351	Overall	0.051	-0.314 to 0.416	.785
3	277	Age >15 y	-0.074	-0.517 to 0.368	.742
2	47	Age 13-15 y	0.536	0.243 to 1.316	.177
2	27	Age ≤12 y	0.392	0.070 to 0.854	.097

Adapted from Benkhadra et al (2017).<sup>12</sup>

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>.

Earlier meta-analyses of glucose monitoring devices for type 1 diabetes tended to combine studies of intermittent glucose monitoring with studies of long-term CGM. Several reported separate subgroup analyses for long-term CGM. A Cochrane review by Langendam et al (2012) assessed CGM in type 1 diabetes in adults and children included RCTs; it compared CGM with conventional SMBG.<sup>8</sup> In pooled analysis (6 studies; n=963 patients) of studies of long-term CGM, the average decline in HbA1c levels 6 months after baseline was statistically significantly larger for CGM users than for SMBG users (mean difference [MD], -0.2%; 95% confidence interval [CI], -

0.4% to -0.1%), but there was no difference in the decline in HbA1c levels at 12 months (1 study, n=154 patients; MD, 0.1%; 95% CI, -0.5% to 0.7%). In a meta-analysis of 4 RCTs (n=689 patients), there was no significant difference in the risk of severe hypoglycemia between CGM and SMBG users and the CI for the relative risk was wide (relative risk, 1.05; 95% CI, 0.63 to 1.77), indicating lack of precision in estimating the effect of CGM on hypoglycemia risk. Reviewers were unable to compare the longer-term changes in HbA1c levels or hypoglycemia outcomes for real-time CGM. Trials reporting results by compliance subgroups found larger treatment effects in highly compliant patients.

A systematic review by Wojciechowski et al (2011) evaluating CGM included RCTs conducted in adults and children with type 1 diabetes.<sup>10</sup> Reviewers selected studies having a minimum of 12 weeks of follow-up and requiring patients to be on intensive insulin regimens. Studies compared CGM with SMBG; there was no restriction on the type of CGM device but CGM readings had to be used to adjust insulin dose or modify diet. Fourteen RCTs met the eligibility criteria. Study durations ranged from 3 to 6 months. Baseline mean HbA1c levels ranged from 6.4% to 10%. Five included studies found a statistically significant decrease in HbA1c levels favoring CGM, while 9 did not. In a pooled analysis, there was a statistically significant reduction in HbA1c levels with CGM compared with SMBG (weighted mean difference [WMD], -0.26%; 95% CI, -0.34% to -0.19%). For the subgroup of 7 studies that reported on long-term CGM, this difference was statistically significant (WMD = -0.26; 95% CI, -0.34 to -0.18). In a subgroup analysis by age, there were significant reductions in HbA1c levels with CGM in 5 studies of adults (WMD= -0.33; 95% CI, -0.46 to -0.20) and in 8 studies with children and/or adolescents (WMD= -0.25; 95% CI, -0.43 to -0.08). Four of the studies provided data on the frequency of hypoglycemic episodes. Pooled results showed a significant reduction in hypoglycemic events for CGM versus SMBG (standardized mean difference, -0.32; 95% CI, -0.52 to -0.13). In 5 studies reporting the percentage of patients with severe hypoglycemic episodes, there were no differences in the percentages of patients with severe hypoglycemic episodes using CGM and SMBG.

### Randomized Controlled Trials

Recent RCTs not included in the meta-analyses above are described next and in Tables 3 and 4. HbA1c, blood glucose, event rates, and patient reported outcomes were assessed at 6 months. None of the studies were blinded. The studies had a large number of pre-specified secondary endpoints, and analyses took into consideration the statistical impact of multiple comparisons.

Two, 2017 RCTs evaluated long-term CGM in patients with type 1 diabetes treated with multiple daily insulin injections. Both trials used the Dexcom G4 CGM device. Lind et al (2017) reported on a crossover study with 142 adults ages 18 and older who had baseline HbA1c levels of 7.5% or higher (mean baseline HbA1c level, »8.5%).<sup>13</sup> Enrolled patients underwent 26-week treatment periods with a CGM device and conventional therapy using SMBG, in random order. There was a 17-week washout period between intervention phases. The primary endpoint was the difference in HbA1c levels at the end of each treatment period. Mean HbA1c levels were 7.9% during CGM use and 8.4% during conventional therapy (MD = -0.4%; p<.01). Treatment satisfaction (measured by the Diabetes Treatment Satisfaction Questionnaire) was significantly higher in the CGM phase than in the conventional treatment phase (p<.001). There was 1 (0.7%) severe hypoglycemic event during the CGM phase and 5 (3.5%) events during conventional therapy. The percentage of time with hypoglycemia (<70 mmol/L) was 2.8% during CGM treatment and 4.8% during conventional therapy.

In the second study, Beck et al (2017) randomized 158 patients on a 2:1 basis to 24 weeks of CGM (n=105) or usual care (n=53).<sup>14</sup> The primary outcome (change in HbA1c levels at 24 weeks) was 1.0% in the CGM group and 0.4% in the usual care group (p<.001), with a between-group difference of 0.6%. Prespecified secondary outcomes on the proportion of patients below a glycemic threshold at 24 weeks also favored the CGM group. The proportion of patients with HbA1c levels less than 7.0% was 18 (18%) in the CGM group and 2 (4%) in the control group (p=.01). Prespecified secondary outcomes related to hypoglycemia also differed significantly between groups, favoring the CGM group. Comparable numbers for time spent at less than 50

mg/dL were 6 minutes per day in the CGM group and 20 minutes per day in the usual care group ( $p=.001$ ). The median change in the rate per 24 hours of hypoglycemia events lasting at least 20 minutes at less than 3.0 mmol/L (54 mg/dL) fell by 30% from 0.23 at baseline to 0.16 during follow-up in the CGM group but was practically unchanged (0.31 at baseline and 0.30 at follow-up) in the usual care group ( $p=.03$ ).<sup>15</sup> Quality of life measures assessing overall well-being (World Health Organization Well-Being Index), health status (EQ-5D-5L), diabetes distress (Diabetes Distress Scale), hypoglycemic fear (worry subscale of the Hypoglycemia Fear Survey), and hypoglycemic confidence (Hypoglycemic Confidence Scale) have also been reported.<sup>16</sup> There were no significant differences between CGM and usual care in changes in well-being, health status, or hypoglycemic fear. The CGM group demonstrated a greater increase in hypoglycemic confidence ( $p=.01$ ) and a greater decrease in diabetes distress ( $p=.01$ ) than the usual care group.

Two RCTs were published in 2020 that assessed CGM with a Dexcom G5 in adolescents and young adults (Laffel et al, 2020)<sup>17</sup>, and in older adults (Pratley et al, 2020)<sup>18</sup>. Both studies found modest but statistically significant differences in HbA1c between patients who used the CGM devices compared to the control arm at follow-up. Secondary measures of HbA1c and blood glucose were mostly better in the CGM arm. Patient-reported outcome measures were not significantly different between the groups, except that glucose monitoring satisfaction was higher in the adolescents and young adults who used CGM. With the newer technology, patients were able to use a smartphone app to monitor glucose levels.

**Table 3. Summary of Key RCT Characteristics in Patients with Type 1 Diabetes**

Study; Trial	Countries	Sites	Dates	Participants	Interventions	
					CGM	SMBG
Beck et al (2017) <sup>14</sup> . DIAMOND				Adults aged 25 or older with baseline HbA1c levels between 7.5% and 10%	Dexcom G4 (n=105)	Usual care (n=53)
Laffel et al (2020) <sup>17</sup> .	US	14	2018-2019	Adolescents and young adults age 14 to 24 years with HbA1c 7.5% to 10.9% with multiple daily insulin injections or an insulin pump	Dexcom G5, with training on use and a smartphone app and 2 calibration BG per day (n=74)	Fingerstick blood glucose meter checks at least 4 times daily (n=79)
Pratley et al (2020) <sup>18</sup> . (WISDM)	US	22	1993-2012	Older adults $\geq 60$ years of age with HbA1c < 10.0% with multiple daily insulin injections or an insulin pump	Dexcom G5 with training on use and 2 calibration BG per day (n=103)	Fingerstick blood glucose meter checks at least 4 times daily (n=100)

BG: blood glucose; CGM: continuous glucose monitoring; HbA1c: hemoglobin A1C; RCT: randomized controlled trial; SMBG: self-monitored blood glucose; WISDM: Wireless Innovation for Seniors With Diabetes Mellitus.

**Table 4. Summary of Key RCT Results**

Study	HbA1c	HbA1c	Blood Glucose (SD) mg/dL	Hypoglycemic Episodes	Patient Reported Outcomes
Beck et al (2017) <sup>14</sup> . DIAMOND	<i>Change from Baseline</i>	<i>Proportion &lt;7.0%</i>		<i>Minutes per day &lt;70 mg/dL</i>	
CGM	1.0%	18 (18%)		43	
SMBG	0.4%	2 (4%)		80	
Diff (95% CI)	0.6%				

Study	HbA1c	HbA1c	Blood Glucose (SD) mg/dL	Hypoglycemic Episodes	Patient Reported Outcomes	
P	<.001	.01		0.002		
Laffel et al (2020) <sup>1Z</sup>	<i>Change from Baseline</i>	<i>Percent with Reduction of 0.5%</i>	<i>Mean (SD)</i>	<i>Per Week</i>	<i>PAD-PS Survey</i>	<i>Glucose Monitoring Satisfaction</i>
CGM	-0.4 (1.0)	44%	199 (36)	1.4 (0.4 to 2.6)		
SMBG	0.1 (0.8)	21%	217 (35)	1.7 (1.0 to 3.1)		
Diff (95% CI)	-0.37 (-0.66 to -0.08)	23% (7% to 37%)	-14.3 (-23.6 to -5.1)	-0.3 (-0.7 to 0.1)	-0.1 (-3.0, 4.0)	0.27 (0.06, 0.54)
P	.01	.005	.003	.11	.73	.003
Pratley et al (2020) <sup>1B</sup> (WISDM)	<i>At follow-up</i>	<i>Percentage of time glucose values &lt; 70 mg/dL</i>		<i>Per week</i>	<i>Quality of life</i>	<i>Hypoglycemia Awareness</i>
CGM	7.2 (0.9)	2.7%	162 (23)	0.8 (0.3-2.2)		
SMBG	7.4 (0.9)	4.9%	171 (30)	1.8 (0.7-4.0)		
Diff (95% CI)	-0.3 (-0.4 to -0.1)	-1.9% (-2.8 to -1.1)	-7.7 (-13.1 to -2.4)	-0.9 (-1.3 to -0.5)		
P		<.001	.005	<.001	NS	NS
Summary <sup>2</sup>	Range					

CI: confidence interval; CGM: continuous glucose monitor; HbA1c: hemoglobin A1c; NS: not significant; PAD-PS; Problem Areas in Diabetes-Pediatric Survey; RCT: randomized controlled trial; SD: standard deviation; SMBG: self monitored blood glucose; WISDM: Wireless Innovation for Seniors With Diabetes Mellitus

### Pregnant Women

One trial of real-time CGM in pregnant women with type 1 diabetes has been reported. Study characteristics, results, and gaps are summarized here and in Tables 5 to 8. Feig et al (2017) reported results of 2 multicenter RCTs in women ages 18 to 40 with type 1 diabetes who were receiving intensive insulin therapy and who were either pregnant ( $\leq 13$  weeks and 6 days of gestation) or planning a pregnancy.<sup>12</sup> The trial enrolling pregnant women is reviewed here. Women were eligible if they had a singleton pregnancy and HbA1c levels between 6.5% and 10.0%. The trial was conducted at 31 hospitals in North America and Europe. Women were randomized to CGM (Guardian REAL-Time or MiniMed Minilink system) plus capillary glucose monitoring or capillary glucose monitoring alone. Women in the CGM group were instructed to use the devices daily. Women in the control group continued their usual method of capillary glucose monitoring. The target glucose range was 3.5 to 7.8 mmol/L and target HbA1c levels were 6.5% or less in both groups. The primary outcome was the difference in change in HbA1c levels from randomization to 34 weeks of gestation. The proportion of completed scheduled study visits was high in both groups; however, participants using CGM had more unscheduled contacts, which were attributed both to sensor issues and to sensor-related diabetes management issues. The median frequency of CGM use was 6.1 days per week (interquartile range, 4.0 to 6.8 d/wk) and 70% of pregnant participants used CGM for more than 75% of the time. The between-group difference in the change in HbA1c levels from baseline to 34 weeks of gestation was statistically significant favoring CGM (MD = -0.19%; 95% CI, -0.34 to -0.03; p=.02). Women in the CGM group spent an increased percentage of time in the recommended glucose control target range at 34 weeks of gestation (68% vs 61%, p=.003). There were no between-group differences in maternal hypoglycemia, gestational weight gain, or total daily insulin dose. A smaller proportion of infants of mothers in the CGM group were large-for-gestational-age (odds ratio [OR], 0.51; 95% CI, 0.28 to 0.90; p=.02). In addition, for infants of mothers in the CGM group, there were fewer neonatal intensive care admissions lasting more than 24 hours (OR=0.48; 95% CI, 0.26 to 0.86; p=.02), fewer incidences of neonatal hypoglycemia

requiring treatment with intravenous dextrose (OR=0.45, 0.22 to 0.89; p=.025), and reduced total hospital length stay (3.1 days vs 4.0 days; p=.0091). Skin reactions occurred in 49 (48%) of 103 CGM participants and 8 (8%) of 104 control participants.

**Table 5. RCT Characteristics for Real-Time CGM in Pregnant Women With Type 1 Diabetes**

Study; Registration	Countries	Sites	Dates	Participants	Interventions	
					Active	Comparator
Feig et al (2017) <sup>12</sup> ; NCT01788527	Canada, England, Scotland, Spain, Italy, Ireland, U.S.	31	2013-2016	Pregnant women (<14 wk gestation) with type 1 diabetes receiving intensive insulin therapy with HbA1c levels between 6.5% and 10.0% (mean, 6.9%); mean age, 31 y	CGM (real-time) (n=108)	SMBG (n=107)

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; NCT: national clinical trial; RCT: randomized controlled trial; SMBG: self-monitored blood glucose.

**Table 6. RCT Outcomes for Real-Time CGM in Pregnant Women With Type 1 Diabetes**

Study	Infant Large-for-Gestational Age	Gestational Age at Delivery, wk	Severe Hypoglycemia	Caesarean Section	Maternal HbA1c Levels: Change From Baseline to 34 Wk of Gestation	Severe Hypoglycemia
Feig et al (2017) <sup>12</sup>						
n	211	201	200	202	173	214
CGM	53 (53%)	Median, 37.4	15 (15%)	63 (63%)	-0.54	11 (11%)
Control	69 (69%)	Median, 37.3	28 (28%)	74 (73%)	-0.35	12 (12%)
TE (95% CI)	OR=0.51 (0.28 to 0.90)	NR	OR=0.45 (0.22 to 0.89)	NR	-0.19% (-0.34% to -0.03%)	NR
p	.02	.50	.025	.18	.02	1.0

Values are n or n (%) or as otherwise indicated.

CI: confidence interval; CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; NR: not reported; OR: odds ratio; RCT: randomized controlled trial; TE: treatment effect.

The purpose of the limitations tables (see Tables 7 and 8 ) is to display notable limitations identified in each study. This information is synthesized as a summary of the body of evidence following each table and provides the conclusions on the sufficiency of evidence supporting the position statement.

**Table 7. Study Relevance Limitations of RCTs for Real-Time CGM in Pregnant Women With Type 1 Diabetes**

Study	Population <sup>a</sup>	Intervention <sup>b</sup>	Comparator <sup>c</sup>	Outcomes <sup>d</sup>	Follow-Up <sup>e</sup>
Feig et al (2017) <sup>12</sup>	4. Run-in period requirement may have biased selection to highly compliant participants	3. More unscheduled contacts in CGM group	3. More unscheduled contacts in CGM group		

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CGM: continuous glucose monitoring; RCT: randomized controlled trial.

<sup>a</sup> Population key: 1. Intended use population unclear; 2. Clinical context is unclear; 3. Study population is unclear; 4. Study population not representative of intended use.

<sup>b</sup> Intervention key: 1. Not clearly defined; 2. Version used unclear; 3. Delivery not similar intensity as

comparator; 4. Not the intervention of interest.

<sup>c</sup> Comparator key: 1. Not clearly defined; 2. Not standard or optimal; 3. Delivery not similar intensity as intervention; 4. Not delivered effectively.

<sup>d</sup> Outcomes key: 1. Key health outcomes not addressed; 2. Physiologic measures, not validated surrogates; 3. No CONSORT reporting of harms; 4. Not establish and validated measurements; 5. Clinical significant difference not prespecified; 6. Clinical significant difference not supported.

<sup>e</sup> Follow-Up key: 1. Not sufficient duration for benefit; 2. Not sufficient duration for harms.

**Table 8. Study Design and Conduct Limitations of RCTs for Real-Time CGM in Pregnant Women With Type 1 Diabetes**

Study	Allocation <sup>a</sup>	Blinding <sup>b</sup>	Selective Reporting <sup>c</sup>	Data Completeness <sup>d</sup>	Power <sup>e</sup>	Statistical <sup>f</sup>
Feig et al (2017) <sup>12</sup>		1. Not blinded; chance of bias in clinical management				3, 4. Treatment effects and confidence intervals not calculated for some outcomes

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CGM: continuous glucose monitoring; RCT: randomized controlled trial.

<sup>a</sup> Allocation key: 1. Participants not randomly allocated; 2. Allocation not concealed; 3. Allocation concealment unclear; 4. Inadequate control for selection bias.

<sup>b</sup> Blinding key: 1. Not blinded to treatment assignment; 2. Not blinded outcome assessment; 3. Outcome assessed by treating physician.

<sup>c</sup> Selective Reporting key: 1. Not registered; 2. Evidence of selective reporting; 3. Evidence of selective publication.

<sup>d</sup> Data Completeness key: 1. High loss to follow-up or missing data; 2. Inadequate handling of missing data; 3. High number of crossovers; 4. Inadequate handling of crossovers; 5. Inappropriate exclusions; 6. Not intent to treat analysis (per protocol for noninferiority trials).

<sup>e</sup> Power key: 1. Power calculations not reported; 2. Power not calculated for primary outcome; 3. Power not based on clinically important difference.

<sup>f</sup> Statistical key: 1. Analysis is not appropriate for outcome type: (a) continuous; (b) binary; (c) time to event; 2. Analysis is not appropriate for multiple observations per patient; 3. Confidence intervals and/or p values not reported; 4. Comparative treatment effects not calculated.

### Continuous Glucose Monitoring Implanted Device for Long-Term Use

The Eversense Continuous Glucose Monitoring System is implanted in the subcutaneous skin layer and provides continuous glucose measurements over a 40 to 400 mg/dL range. The system provides real-time glucose values, glucose trends, and alerts for hypoglycemia and hyperglycemia and low glucose through a mobile application installed on a compatible mobile device platform. The Eversense CGM System is a prescription device indicated for use in adults (age 18 and older) with diabetes for up to 90 days. The device was initially approved as an adjunctive glucose monitoring device to complement information obtained from standard home blood glucose monitoring devices. Prescribing providers are required to participate in insertion and removal training certification.

Data from 3 nonrandomized prospective studies (PRECISE, PRECISE II, AND PRECISION) were provided to the U.S. Food and Drug Administration (FDA) for the initial approval of Eversense as an adjunctive device.<sup>20,21</sup> Expanded approval was granted in June 2019 and Eversense is now approved as a device to replace fingerstick blood glucose measurements for diabetes treatment decisions.<sup>22</sup> Historical data from the system can be interpreted to aid in providing therapy adjustments. No new clinical studies were conducted to support the change in the indications for the device. The sponsor had previously performed clinical studies to establish the clinical measurement performance characteristics of the device, including accuracy across the claimed measuring range (40 to 400 mg/dL glucose), precision, claimed calibration frequency (every 12 hours), the wear period for the sensor (90 days), and performance of the alerts and notifications. This same clinical study information was used to support what the FDA considered a reasonable assurance of safety and effectiveness of the device for the replacement of

fingerstick blood glucose monitoring for diabetes treatment decisions. As a condition of approval, the sponsor is required to conduct a post-approval-study to evaluate the safety and effectiveness of diabetes management with the Eversense CGM System non-adjunctively compared to self-monitoring of blood glucose using a blood glucose meter in participants with either Type 1 or Type 2 diabetes.<sup>22</sup> The study design is a non-blinded, prospective, multi-center, single-arm longitudinal cohort study. Subjects will serve as their own control, with baseline SMBG use to manage their diabetes for the first 6 months of the study followed by the use of the CGM nonadjunctively for the next 6 months. Total follow-up duration is 12 months. Approximately 925 subjects will be screened to achieve an enrollment such that approximately 740 subjects will be available for analysis at the end of the study. The investigation will include both clinic visits and home use of the device.

Three post-marketing registry studies of the Eversense device have been published (Tables 9 and 10). Sanchez et al (2019) reported glucometric and safety data on the first 205 patients in the U.S. to use the Eversense device for at least 90 days.<sup>23</sup> Of the 205 patients, 62.9% reported having type 1 diabetes, 8.8% type 2 diabetes, and 28.3% were unreported; results were not reported separately by diabetes type. Diess et al (2019) reported safety outcomes for 3023 patients from 534 sites in Europe and South Africa who had used the device for 6 months or longer.<sup>24</sup> There were no serious adverse events, and the most commonly reported adverse events were sensor site infection and skin irritation. Tweden et al (2019) reported accuracy and safety data from 945 patients in Europe and South Africa who used either the 90-day or 180 day Eversense system for 4 insertion-removal cycles.<sup>25</sup> The percentage of patients using the 180-day system increased from cycle 1 to 4 as the device became more widely available (9%, 39%, 68% and 88% in cycles 1 to 4). There was no evidence of degradation of performance of the device over repeated insertion/removal cycles. Adverse events were not otherwise reported.

**Table 9. Summary of Key Nonrandomized Trials: Implanted CGM Study Characteristics**

Study	Study Type	Country	Dates	Participants	Test/Treatment	Follow-Up
Deiss et al (2019) <sup>24</sup>	Prospective Single-arm Unblinded Postmarketing registry	Europe and South Africa	2016-2018	Adults (≥ 18 years) with T1D or T2D (% not reported) Consecutive patients who reached 4 sensor insertion/removal cycles Total N=3023; 6 months of use (N=969), 1 year of use (N=173)	Implanted CGM Single sensor (90-day or 180 days)	Up to 1 year
Sanchez et al (2019) <sup>23</sup>	Prospective Single-arm Unblinded Postmarketing registry	United States	2018-2019	Consecutive patients who reached a 90-day wear period of the device (62.9% T1D, 8.8% T2D, 28.3% unreported) (N=205)	Implanted CGM	90 days
Tweden et al (2019) <sup>25</sup>	Prospective Single-arm Unblinded Postmarketing registry	Europe and South Africa	2016-2019	Adult patients with T1D or T2D (% not known) for whom the Eversense CGM System was prescribed and inserted by their health care provider across approximately 1000 centers in Europe and South Africa (N=945)	Implanted CGM 90 day system or 180 day system	4 insertion-removal cycles

CGM: continuous glucose monitoring; T1D: type 1 diabetes; T2D: type 2 diabetes.

**Table 10. Summary of Key Nonrandomized Trials: Implanted CGM Study Results**

Study	Efficacy Results	Adverse Events
Deiss et al (2019) <sup>24</sup>		N=3023

Study	Efficacy Results	Adverse Events
<b>Efficacy Outcomes</b>		
	NR (safety only)	133 adverse events (85 procedure-related, 22 device-related, 6 drug-related, 4 device/procedure related; 16 not related) No related serious adverse events through 4 insertion/removal cycles. infection (n=29 patients); adhesive patch irritation (n=20 patients); unsuccessful first removal attempt (n=23 patients)
Sanchez et al (2019) <sup>23</sup> .	N=205	N=205
MARD (glucose range 40-400 mg/dl)	11.2% (SD 11.3%, median 8.2%).	
Mean SG (mg/dL)	161.8 Median 157.2 (IQR 138.4 to 178.9)	
% SG values in hypoglycemia (<54 mg/dL), 24-hour period	1.2% (18.0 minutes)	
% SG values in hypoglycemia (<54 mg/dL), nighttime	1.7%	10 (5%) transient skin irritation, redness, and/or swelling, 4 (2%) mild infection, 3 (1.5%) hypoglycemia that was self-treated, 4 (2%) failure to remove the sensor on the first attempt, and 5 (2.5%) skin irritation due to the adhesive
TIR, 24-hour period	62.3% (~15 hours)	
TIR, nighttime	61.8%	
Time in mild hyperglycemia, 24-hour period	21.9%	
Time in mild hyperglycemia, nighttime	21.5%	
Time in significant hyperglycemia, 24-hour period	11.6%	
Time in significant hyperglycemia, nighttime	12.1%	
Tweden et al (2019) <sup>25</sup> .		
MARD (glucose range 40-400 mg/dl)	Mean 11.5% to 11.9% during each sensor cycle	
Mean SG (mg/dL)	156.5 to 158.2 mg/dL across 4 sensor cycles	
% SG values in significant hypoglycemia (<54 mg/dL), 24-hour period	1.1% to 1.3% (16 to 19 minutes)	No evidence of degradation of performance from the repeated insertion and removal procedures occurring in approximately the same subcutaneous tissue of the body. Adverse events otherwise not reported.
% SG values in significant hypoglycemia (<70 mg/dL), 24-hour period	4.6% to 5.0% (66 to 72 minutes)	
TIR, 24-hour period	63.2% to 64.5% (910 to 929 minutes)	
Time in hyperglycemia (>180-250 mg/dL), 24-hour period	22.8% to 23.2% (328 to 334 minutes)	
Time in significant hyperglycemia (>250 mg/dL), 24-hour period	8.1% to 8.8% (117 to 127 minutes)	

CGM: continuous glucose monitoring; CI: confidence interval; IQR: interquartile range; MARD: mean absolute relative difference; NR: not reported; SD: standard deviation; SG: sensor glucose; TIR: time in range

### Section Summary: Continuous Glucose Monitoring Devices for Long-Term Use in Type 1 Diabetes

Numerous RCTs and several systematic reviews of RCTs have evaluated CGM in patients with type 1 diabetes. A 2017 individual patient data analysis, using data from 11 RCTs, found that reductions in HbA1c levels were significantly greater with real-time CGM compared with a control intervention. In addition, a 2012 meta-analysis of 6 RCTs found a significantly larger decline in HbA1c levels at 6 months in the CGM group than the SMBG group. There are few studies beyond 6 months. Two recent RCTs in patients who used multiple daily insulin injections and were highly compliant with CGM devices during run-in phases found that CGM was

associated with a larger reduction in HbA1c levels than previous studies. Reductions were 0.4% and 0.6%, respectively, compared with approximately 0.2% to 0.3% in previous analyses. One of the 2 RCTs prespecified hypoglycemia-related outcomes and time spent in hypoglycemia were significantly lower in the CGM group.

One RCT in pregnant women with type 1 diabetes (n=215) has compared CGM with SMBG. Adherence was high in the CGM group. The difference in the change in HbA1c levels from baseline to 34 weeks of gestation was statistically significant favoring CGM, and women in the CGM group spent an increased percentage of time in the recommended glucose control target range at 34 weeks of gestation. There were no between-group differences in maternal hypoglycemia, gestational weight gain, or total daily insulin dose. A smaller proportion of infants of mothers in the CGM group were large for gestational age, had neonatal intensive care admissions lasting more than 24 hours, and had neonatal hypoglycemia requiring treatment. The total hospital length of stay was shorter by almost 1 day in the CGM group.

Three nonrandomized prospective studies and 3 postmarketing registry studies assessed the accuracy and safety of an implanted glucose monitoring system that provides continuous glucose monitoring for up to 4 insertion/removal cycles as an adjunct to home glucose monitoring devices. Accuracy measures included the mean absolute relative difference between paired samples from the implanted device and a reference standard blood glucose measurement. The accuracy tended to be lower in hypoglycemic ranges. Limitations on the evidence include lack of differentiation in outcomes for type 1 diabetes versus type 2 diabetes and variability in reporting of trends in secondary glycemic measures. The initial approval of the device has been expanded to allow the device to be used for glucose management decision making. The same clinical study information was used to support what the FDA considered a reasonable assurance of safety and effectiveness of the device for the replacement of fingerstick blood glucose monitoring for diabetes treatment decisions. As a condition of approval, the sponsor is required to conduct an additional post-approval-study.

### **Continuous Glucose Monitoring Devices for Short-Term Use in Type 1 Diabetes Clinical Context and Therapy Purpose**

The purpose of the short-term use of CGM devices is to provide a testing option that is an alternative to or an improvement on existing testing used in the management of individuals with type 1 diabetes.

The question addressed in this evidence review is: Does the short-term use of a CGM device improve the net health outcome for individuals with type 1 diabetes?

The following PICO was used to select literature to inform this review.

#### ***Populations***

The relevant population of interest is individuals with type 1 diabetes. All individuals with type 1 diabetes require engagement in a comprehensive self-management and clinical assessment program that includes assessment of blood glucose control. Individuals with type 1 diabetes may have poorly controlled diabetes, despite current use of best practices, including situations such as unexplained hypoglycemic episodes, hypoglycemic unawareness, suspected postprandial hyperglycemia, and recurrent diabetic ketoacidosis. In addition, individuals with type 1 diabetes may need to determine basal insulin levels prior to insulin pump initiation.

#### ***Interventions***

The testing being considered is the short-term use of a CGM device to assess blood glucose levels as part of optimal diabetes management. Short-term use is generally for 72 hours. However, reports of use range from 3 to 30 days.

### ***Comparators***

The following practice is currently being used to measure glucose levels: capillary blood sampling (finger stick) for SMBG. Standard treatment for patients with type 1 diabetes includes injection of long-acting basal insulin plus MDI of rapid-acting insulin boluses as required for meal intake. Activity level may require patients need to modify the timing and dose of insulin administration. Individuals with type 1 diabetes may also use an insulin pump either for initial treatment or convert to pump use after a period of MDI. Individuals are required to check their blood glucose before making preprandial insulin calculations, in response to symptoms of hypoglycemia or related to activity-related insulin adjustments

### ***Outcomes***

For short-term use of CGM, the general outcomes of interest include time in range (generally glucose of 70 to 180 mg/dl), frequency and time spent in hypoglycemia and, frequency and time spent in hyperglycemia for the duration of the monitoring. Repeat CGM may be necessary to assess the impact of changes in management.

### **Study Selection**

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs;
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess long-term outcomes and adverse effects, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

### **Review of Evidence**

#### **Systematic Reviews**

Meta-analyses of glucose monitoring devices for type 1 diabetes tend to combine studies of short-term glucose monitoring with studies of long-term CGM. For this body of evidence, there is variability in the definitions of intermittent monitoring and the specific monitoring protocols used. Also, many of the trials of short-term monitoring have included additional interventions to optimize glucose control (e.g., education, lifestyle modifications).

Two meta-analyses were identified that reported separate subgroup analyses for intermittent monitoring. In a Cochrane review by Langendam et al (2012), 4 studies (total n=216 patients) compared real-time intermittent glucose monitoring systems with SMBG, and the pooled effect estimate for change in HbA1c levels at 3 months was not statistically significant (MD change, -0.18; 95% CI, -0.42 to 0.05).<sup>8</sup> The meta-analysis by Wojciechowski et al (2011), which assessed RCTs on CGM (described previously), also included a separate analysis of 8 RCTs of intermittent monitoring.<sup>10</sup> On pooled analysis, there was a statistically significant reduction in HbA1c levels with intermittent glucose monitoring compared with SMBG (WMD= -0.26; 95% CI, -0.45 to -0.06).

#### **Randomized Controlled Trials**

The largest RCT was the Management of Insulin-Treated Diabetes Mellitus (MITRE) trial, published by Newman et al (2009); it evaluated whether the use of the additional information provided by minimally invasive glucose monitors improved glucose control in patients with poorly controlled insulin-requiring diabetes.<sup>26</sup> This 4-arm RCT was conducted at secondary care diabetes clinics in 4 hospitals in England. This trial enrolled 404 people over the age of 18 years, with insulin-treated diabetes (types 1 or 2) for at least 6 months, who were receiving 2 or more injections of insulin daily. Most (57%) participants had type 1 diabetes (41% had type 2 diabetes, 2% were classified as "other"). Participants had to have 2 HbA1c values of at least 7.5% in the 15 months before trial entry and were randomized to 1 of 4 groups. Two groups received minimally invasive glucose monitoring devices (GlucoWatch Biographer or MiniMed Continuous Glucose Monitoring System [CGMS]). Intermittent glucose monitoring was used (i.e., monitoring was performed over several days at various points in the trial). These groups were compared with an attention control group

(standard treatment with nurse feedback sessions at the same frequency as those in the device groups) and a standard control group (reflecting common practice in the clinical management of diabetes). Changes in HbA1c levels from baseline to 3, 6, 12, and 18 months were the primary indicator of short- to long-term efficacy. At 18 months, all groups demonstrated a decline in HbA1c levels from baseline. Mean percentage changes in HbA1c levels were -1.4% for the GlucoWatch group, -4.2% for the CGMS group, -5.1% for the attention control group, and -4.9% for the standard care control group. In the intention-to-treat analysis, no significant differences were found between any groups at any assessment times. There was no evidence that the additional information provided by the devices changed the number or nature of treatment recommendations offered by the nurses. Use and acceptability indicated a decline for both devices, which was most marked in the GlucoWatch group by 18 months (20% still using GlucoWatch vs 57% still using the CGMS). In this trial of unselected patients, glucose monitoring (CGMS on an intermittent basis) did not lead to improved clinical outcomes.

## Pregnant Women

### Systematic Reviews

Voormolen et al (2013) published a systematic review of the literature on CGM during pregnancy.<sup>27</sup> They identified 11 relevant studies (total n=534 women). Two were RCTs, one of which was the largest of the studies (n=154). Seven studies used CGMs that did not have data available in real-time; the remaining 4 studies used real-time CGM. Reviewers did not pool study findings; they concluded that the evidence was limited to the efficacy of CGM during pregnancy. The published RCTs are described next.

### Randomized Controlled Trials

Two RCTs of intermittent glucose monitoring in pregnant women with type 1 or type 2 diabetes are summarized in Tables 11 to 14 and the following paragraphs. While both trials included a mix of women with type 1 and type 2 diabetes, most women had type 1 diabetes in both trials, so the trials are reviewed in this section.

Secher et al (2013) randomized 154 women with type 1 (n=123) and type 2 (n=31) diabetes to real-time CGM in addition to routine pregnancy care (n=79) or routine pregnancy care alone (n=75).<sup>28</sup> Patients in the CGM group were instructed to use the CGM device for 6 days before each of 5 study visits and were encouraged to use the devices continuously; 64% of participants used the devices per-protocol. Participants in both groups were instructed to perform 8 daily self-monitored plasma glucose measurements for 6 days before each visit. Baseline mean HbA1c levels were 6.6% in the CGM group and 6.8% in the routine care group. The 154 pregnancies resulted in 149 live births and 5 miscarriages. The prevalence of large-for-gestational-age infants (at least 90th percentile), the primary study outcome, was 45% in the CGM group and 34% in the routine care group. The difference between groups was not statistically significant (p=.19). Also, no statistically significant differences were found between groups for secondary outcomes, including the prevalence of preterm delivery and the prevalence of severe neonatal hypoglycemia. Women in this trial had low baseline HbA1c levels, which might explain the lack of impact of CGM on outcomes. Other factors potentially contributing to the negative findings included the intensive SMBG routine in both groups and the relatively low compliance rate in the CGM group.

Murphy et al (2008) in the U.K. randomized 71 pregnant women with type 1 (n=46) and type 2 (n=25) diabetes to CGM or usual care.<sup>29</sup> The intervention consisted of up to 7 days of CGM at intervals of 4 to 6 weeks between 8 weeks and 32 weeks of gestation. Neither participants nor physicians had access to the measurements during sensor use; data were reviewed at study visits. In addition to CGM, the women were advised to measure blood glucose levels at least 7 times a day. Baseline HbA1c levels were 7.2% in the CGM group and 7.4% in the usual care group. The primary study outcome was maternal glycemic control during the second and third trimesters. Eighty percent of women in the CGM group wore the monitor at least once per trimester. Mean HbA1c levels were consistently lower in the intervention arm, but differences between groups were statistically significant only at week 36. For example, between 28 weeks

and 32 weeks of gestation, mean HbA1c levels were 6.1% in the CGM group and 6.4% in the usual care group (p=.10). The prevalence of large-for-gestational-age infants (at least 90th percentile) was a secondary outcome. Thirteen (35%) of 37 infants in the CGM group were large-for-gestational age compared with 18 (60%) of 30 in the usual care group. The odds for reduced risk of a large-for-gestational-age infant with CGM was 0.36 (95% CI, 0.13 to 0.98; p=.05).

**Table 11. RCT Characteristics for Short-Term Continuous Glucose Monitoring in Pregnant Women With Type 1 Diabetes**

Study; Registration	Countries	Sites	Dates	Participants	Interventions	
					Active	Comparator
Secher et al (2013) <sup>28</sup> ; NCT00994357	Denmark	1	2009-2011	Pregnant women with type 1 (80%) or type 2 (20%) diabetes; mean gestational age, <14 wk); median HbA1c level, 6.7%; median age, 32 y	CGM (for 6 d before each study visits; encouraged to used continuously) plus SOC (n=79)	SOC (n=75)
Murphy et al (2008) <sup>29</sup> ; ISRCTN84461581	U.K.	2	2003-2006	Pregnant women with type 1 (65%) or type 2 (35%) diabetes; mean gestational age, 9.2 wk; mean HbA1c level, 7.3%; mean age, 31 y	CGM (up to 7 d of CGM at intervals of 4-6 wk) plus SOC (n=38)	SOC (n=33)

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; NCT: national clinical trial; RCT: randomized controlled trial; SOC: standard of care.

**Table 12. RCT Results for Short-Term Continuous Glucose Monitoring in Pregnant Women With Type 1 Diabetes**

Study	Infant					Maternal HbA1c Levels at 36 Weeks of Gestation <sup>a</sup>	Severe Hypoglycemia
		Large-for-Gestational Age	Gestational Age at Delivery	Severe Hypoglycemia	Caesarean Section		
<b>Days</b>							
Secher et al (2013) <sup>28</sup>							
n	154	154	145	154		154	
CGM	34 (45%)	Median, 263	9 (13%)	28 (37%)	Median, 6.0%		16%
Control	25 (34%)	Median, 264	10 (14%)	33 (45%)	Median, 6.1%		16%
TE (95% CI)	NR	NR	NR	NR	NR		NR
p	.19	.14	.88	.30	.63		.91
<b>Weeks</b>							
Murphy et al (2008) <sup>29</sup>							
n	71	71	68	69	71		NR
CGM	13 (35%)	Mean, 37.6	3 (8%)	27 (71%)	Mean, 5.8%		
Control	18 (60%)	Mean, 37.5	5 (17%)	21 (61%)	Mean, 6.4%		
TE (95% CI)	OR=0.36 (0.13 to 0.98)	NR	NR	NR	0.6% (CI NR)		
p	.05	.80	.50	.40	.007		

Values are n or n (%) or as otherwise indicated.

CGM: continuous glucose monitoring; CI: confidence interval; HbA1c: hemoglobin A<sub>1c</sub>; NR: not reported; OR: odds ratio; RCT: randomized controlled trial; TE: treatment effect.

<sup>a</sup> N inconsistently reported for HbA1c outcome.

In summary, 2 trials of intermittent glucose monitoring conducted in Europe included pregnant women with type 1 or 2 diabetes, with most having type 1 diabetes. Secher et al (2013) included intermittent, real-time monitoring<sup>28</sup>; Murphy et al (2008)<sup>29</sup> included intermittent, retrospective monitoring with CGM. The intervention started in early pregnancy in these studies; the mean age was in the early-30s and mean baseline HbA1c level was greater than 6.5%. There was no statistically significant difference between CGM and routine care for maternal HbA1c levels at 36 weeks in Secher et al (2013); the difference in HbA1c levels at 36 weeks was about 0.6% ( $p=.007$ ) in Murphy et al (2008). Secher et al (2013) also reported no difference in severe maternal hypoglycemia. The proportion of infants that were large for gestational age (>90th percentile) was higher in the CGM group in Secher et al (2013) although not statistically significantly higher; the difference in large for gestational age was statistically significantly lower for CGM in Murphy et al (2008). The differences in the proportions of infants born via cesarean section, gestational age at delivery, and infants with severe hypoglycemia were not statistically significant in either trial.

Tables 13 and 14 display notable limitations identified in each study.

**Table 13. Study Relevance Limitations of RCTs of Intermittent CGM in Pregnant Women With Type 1 Diabetes**

Study	Population <sup>a</sup>	Intervention <sup>b</sup>	Comparator <sup>c</sup>	Outcomes <sup>d</sup>	Follow-Up <sup>e</sup>
Secher et al (2013) <sup>28</sup> .	4. Study population had relatively low HbA1c levels	4. Only 64% of the participants used devices per protocol			
Murphy et al (2008) <sup>29</sup> .					

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; RCT: randomized controlled trial.

<sup>a</sup> Population key: 1. Intended use population unclear; 2. Clinical context is unclear; 3. Study population is unclear; 4. Study population not representative of intended use.

<sup>b</sup> Intervention key: 1. Not clearly defined; 2. Version used unclear; 3. Delivery not similar intensity as comparator; 4. Not the intervention of interest.

<sup>c</sup> Comparator key: 1. Not clearly defined; 2. Not standard or optimal; 3. Delivery not similar intensity as intervention; 4. Not delivered effectively.

<sup>d</sup> Outcomes key: 1. Key health outcomes not addressed; 2. Physiologic measures, not validated surrogates; 3. No CONSORT reporting of harms; 4. Not establish and validated measurements; 5. Clinical significant difference not prespecified; 6. Clinical significant difference not supported.

<sup>e</sup> Follow-Up key: 1. Not sufficient duration for benefit; 2. Not sufficient duration for harms.

**Table 14. Study Design and Conduct Limitations of RCTs of Short-Term CGM Glucose Monitoring in Pregnant Women With Type 1 Diabetes**

Study	Allocation <sup>a</sup>	Blinding <sup>b</sup>	Selective Reporting <sup>c</sup>	Data Completeness <sup>d</sup>	Power <sup>e</sup>	Statistical <sup>f</sup>
Secher et al (2013) <sup>28</sup> .		1. Not blinded; chance of bias in clinical management				3, 4. Treatment effects and confidence intervals not calculated
Murphy et al (2008) <sup>29</sup> .		1. Not blinded; chance of bias in clinical management				3, 4. Treatment effects and confidence intervals not calculated for some outcomes

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CGM: continuous glucose monitoring; RCT: randomized controlled trial.

<sup>a</sup> Allocation key: 1. Participants not randomly allocated; 2. Allocation not concealed; 3. Allocation

concealment unclear; 4. Inadequate control for selection bias.

<sup>b</sup> Blinding key: 1. Not blinded to treatment assignment; 2. Not blinded outcome assessment; 3. Outcome assessed by treating physician.

<sup>c</sup> Selective Reporting key: 1. Not registered; 2. Evidence of selective reporting; 3. Evidence of selective publication.

<sup>d</sup> Data Completeness key: 1. High loss to follow-up or missing data; 2. Inadequate handling of missing data; 3. High number of crossovers; 4. Inadequate handling of crossovers; 5. Inappropriate exclusions; 6. Not intent to treat analysis (per protocol for noninferiority trials).

<sup>e</sup> Power key: 1. Power calculations not reported; 2. Power not calculated for primary outcome; 3. Power not based on clinically important difference.

<sup>f</sup> Statistical key: 1. Analysis is not appropriate for outcome type: (a) continuous; (b) binary; (c) time to event; 2. Analysis is not appropriate for multiple observations per patient; 3. Confidence intervals and/or p values not reported; 4. Comparative treatment effects not calculated.

### **Section Summary: Glucose Monitoring Devices for Short-Term Use in Type 1 Diabetes**

For short-term monitoring of type 1 diabetes, there are few RCTs and systematic reviews. The evidence for short-term monitoring on glycemic control is mixed, and there was no consistency in HbA1c levels. Some trials have reported improvements in glucose control for the intermittent monitoring group but limitations in this body of evidence preclude conclusions. The definitions of control with short-term CGM use, duration of use and the specific monitoring protocols varied. In some studies, short-term monitoring was part of a larger strategy aimed at optimizing glucose control, and the impact of monitoring cannot be separated from the impact of other interventions. Studies have not shown an advantage for intermittent glucose monitoring in reducing severe hypoglycemia events but the number of events reported is generally small and effect estimates imprecise. The limited duration of use may preclude an assessment of any therapeutic effect. Two RCTs of short-term CGM use for monitoring in pregnancy included women with both type 1 and 2 diabetes, with most having type 1 diabetes. One trial reported a difference in HbA1c levels at 36 weeks; the proportion of infants that were large for gestational age (>90th percentile) favored CGM while the second trial did not. The differences in the proportions of infants born via cesarean section, gestational age at delivery, and infants with severe hypoglycemia were not statistically significant in either study. Limitations of the published evidence preclude determining the effects of the technology on net health outcome.

### **Continuous Glucose Monitoring Devices for Use in Type 2 Diabetes**

There is limited ability to distinguish between long-term and short-term glucose monitoring in the analysis of the data for type 2 diabetes, consistent with the literature.

### **Clinical Context and Therapy Purpose**

The purpose of long-term and short-term CGM devices is to provide a treatment option that is an alternative to or an improvement on existing therapies in individuals with type 2 diabetes.

The question addressed in this evidence review is: Does the use of long-term or short-term CGM devices improve the net health outcome for individuals with type 2 diabetes?

The following PICO was used to select literature to inform this review.

### ***Populations***

The relevant population of interest is individuals with type 2 diabetes. All individuals with type 2 diabetes require engagement in a comprehensive self-management and clinical assessment program that includes assessment of blood glucose control. Some individuals with type 2 diabetes may have poorly controlled diabetes, despite current use of best practices, including situations such as unexplained hypoglycemic episodes, hypoglycemic unawareness, and persistent hyperglycemia and A1C levels above target. In addition, some individuals with type 2 diabetes may need to determine basal insulin levels prior to insulin pump initiation.

**Interventions**

The testing being considered is the use of long-term or short-term CGM devices to assess blood glucose levels as part of optimal diabetes management.

**Comparators**

The following practice is currently being used to measure glucose levels: SMBG (capillary blood sampling (finger stick) using blood glucose meters) and periodic measurement of HbA1c.

**Outcomes**

The general outcomes of interest are change in HbA1c levels, frequency of and time spent in hypoglycemia, frequency and time spent in hyperglycemia, complications of hypoglycemia and hyperglycemia, and QOL. To assess short-term outcomes such as HbA1c levels, a minimum follow-up of 8 to 12 weeks is appropriate. To assess long-term outcomes such as time spent in hypoglycemia, the incidence of hypoglycemic events, complications of hypoglycemia, and QOL, follow-up of 6 months to 1 year would be appropriate.

**Study Selection**

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess long-term outcomes and adverse effects, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

**Review of Evidence****Systematic Reviews**

Trials of type 2 diabetes included in the systematic reviews and meta-analyses is shown in Table 15.

**Table 15. CGM Trials for Type 2 Diabetes Included in Systematic Reviews**

Primary Study	Ida et al (2019) <sup>30</sup>
Ehrhardt et al (2011) <sup>31,a</sup>	●
Cosson et al (2009) <sup>32,b</sup>	●
Allen et al (2008) <sup>33,b</sup>	●
Yoo et al (2008) <sup>34,a</sup>	●
Beck et al (2017) <sup>14,a</sup>	●
Ajjan et al (2016) <sup>35,b</sup>	●
Haak et al ((2017) <sup>36,b</sup>	●

CGM: continuous glucose monitoring.

<sup>a</sup> These studies used real-time CGM (RT-CGM) devices compared to SMBG

<sup>b</sup> These studies used retrospective CGM (r-CGM) devices compared to SMBG

A summary of the characteristics of the systematic review is shown in Table 16. Results are described in Table 17.

**Table 16. Systematic Review Characteristics for CGM in Type 2 Diabetes**

Study	Dates	Trials	Participants	N (Range)	Design	Duration
Ida et al (2019) <sup>30</sup>	1960-2018	7	Adults with T2D	669 (25-224)	RCT	At least 8 wk

CGM: continuous glucose monitoring; RCT: randomized controlled trial; T2D: type 2 diabetes.

**Table 17. Meta-Analytic Results for CGM in Type 2 Diabetes**

Study	Reduction in HbA1c Levels (Mean Difference)	Hypoglycemic Events (Mean Difference)	Diabetes Complications (retinopathy, nephropathy, neuropathy, diabetic foot)	Health-Related Quality of Life
<i>Ida et al (2019)</i> <sup>30</sup>				
Total N	660	285		
PE (95% CI)	-0.42 (-0.70 to -0.13)	-0.35 (-0.59 to -0.10) <sup>a</sup>	NR	Multiple diabetes-specific scales used in each study, therefore, results could not be combined for meta-analyses
p	.004	.0006		
I <sup>2</sup>	64%	0%		

CGM: continuous glucose monitoring; CI: confidence interval; HbA1c: hemoglobin A<sub>1c</sub>; NR: not reported; PE: pooled effect.

### Randomized Controlled Trials

Key RCTs of CGM in adults with type 2 diabetes are summarized in Tables 18 to 21. The largest and most recent studies are also briefly summarized in the following paragraphs. Baseline HbA1c levels were between 7.0% and 12.0% in the RCTs, with participants having a mean baseline age range in the mid-50s and early-60s. Most RCTs used a type of intermittent monitoring; some reported data for patients in real-time while others provided data reviewed only at study visits. No studies reported on follow-up beyond 14 months; thus the effect of CGM on outcomes related to diabetic complications is unknown.

An RCT, Multiple Daily Injections and Continuous Glucose Monitoring in Diabetes (DIAMOND), was reported by Beck et al (2017).<sup>37</sup> DIAMOND was performed at 25 endocrinology practices in North America (22 in the U.S., 3 in Canada) and enrolled adults with type 2 diabetes receiving multiple daily injections of insulin. One-hundred fifty-eight patients were randomized in 2 groups: CGM and usual care (n=79 in each group). Patients compliant during a run-in period were eligible for randomization. Patients in both groups were given a blood glucose meter. Participants in the CGM group were given a Dexcom G4 Platinum CGM System (Dexcom) and instructions on use. Change in HbA1c level from baseline to 24 weeks was the primary outcome. Analyses were adjusted for baseline HbA1c levels and the clinic was performed using intention-to-treat analysis with missing data handling by multiple imputations. At baseline, the mean total daily insulin dose was 1.1 U/kg/d. Week 24 follow-up was completed by 97% of the CGM group and 95% of the control group. Mean CGM use was greater than 6 d/wk at 1 month, 3 months, and 6 months. The adjusted difference in mean change in HbA1c level from baseline to 24 weeks was -0.3% (95% CI, -0.5% to 0.0%; p=.022) favoring CGM. The adjusted difference in the proportion of patients with a relative reduction in HbA1c level of 10% or more was 22% (95% CI, 0% to 42%; p=.028) favoring CGM. There were no events of severe hypoglycemia or diabetic ketoacidosis in either group. The treatment groups did not differ in any of the QOL measures.

Ehrhardt and colleagues published 2 reports (2011, 2012) from an RCT evaluating 100 patients.<sup>38,31</sup> The trial evaluated the intermittent use of a CGM device in adults with type 2 diabetes treated with diet/exercise and/or glycemia-lowering medications but not prandial insulin who had an initial HbA1c level of at least 7% but not more than 12%. The trial compared real-time CGM with the Dexcom device used for 4, 2 week cycles (2 weeks on and 1 week off) with SMBG. The primary efficacy outcome was a mean change in HbA1c levels. Mean HbA1c levels in the CGM group were 8.4% at baseline, 7.4% at 12 weeks, 7.3% at 24 weeks, and 7.7% at 52 weeks. In the SMBG group, these values were 8.2% at baseline, 7.7% at 12 weeks, 7.6% at 24 weeks, and 7.9% at 52 weeks. During the trial, the reduction in HbA1c levels was significantly greater in the CGM group than in the SMBG group (p=.04). After adjusting for potential confounders (e.g., age, sex, baseline therapy, whether the individual started taking insulin during the study), the difference between groups over time remained statistically significant (p<.001).

The investigators also evaluated SMBG results for both groups. The mean proportions of SMBG tests less than 70 mg/dL were 3.6% in the CGM group and 2.5% in the SMBG group (p=.06).

Martens et al (2021) reported results of an RCT comparing real-time CGM with SMBG in 176 patients with poorly controlled type 2 diabetes (HbA1c levels 7.8% to 11.5%) treated with basal insulin without prandial insulin. At 8 months, there was a statistically significantly greater decrease in mean HbA1c in the CGM group (adjusted difference -0.4%; 95% CI -0.8% to -0.1%; p=.02), with 1 hypoglycemic event in each group. Aleppo et al (2021) reported a 6-month follow-up study of 163 patients who had been randomized in this same trial (93.1%). Patients originally randomized to SMBG continued to use SMBG for another 6 months, and the CGM group was randomly reassigned either to continue CGM or discontinue CGM and resume SMBG. In the group that discontinued CGM, mean HbA1c increased from 7.9% at 8 months to 8.2% at 14 months, whereas in the group that continued CGM, mean HbA1c decreased from 8.2% to 8.1%.

**Table 18. Key RCT Characteristics for Continuous Glucose Monitoring in Type 2 Diabetes**

Study; Registration	Countries	Sites	Dates	Participants	Interventions	
					Active	Comparator
Beck et al (2017) (DIAMOND) <sup>37</sup> ; NCT02282397	U.S., Canada	25	2014-2016	Adults with T2D using multiple daily injections of insulin with HbA1c levels 7.5%-10.0% (baseline mean, 8.5%); mean age, 60 y	Real-time CGM (n=79)	SMBG (n=79)
Ehrhardt et al (2011) <sup>31</sup>	U.S.	1	NR	Adults with T2D using oral antidiabetic agents without prandial insulin with HbA1c levels 7.0%-12.0% (baseline mean, 8.3%), mean age, 58 y	Real-time CGM for 4 cycles of 3 wk (n=50)	SMBG (n=50)
Martens et al (2021) <sup>39</sup>	U.S.	15	2018-2019	Adults with T2D treated with basal insulin without prandial insulin. HbA1c levels 7.8%-11.5% (baseline mean, 9.1%); mean age, 57 y	Real-time CGM (n=116)	SMBG (n=59)

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; NCT: national clinical trial; NR: not reported; RCT: randomized controlled trial; SMBG: self-monitored blood glucose; T2D: type 2 diabetes.

**Table 19. Key RCT Outcomes for Glucose Monitoring in Type 2 Diabetes**

Study	Reduction in HbA1c Levels (Mean Range), %	HbA1c Level <7.0%, n (%)	Relative Reduction in HbA1c Level ≥10%, n (%)	Hypoglycemic or Ketoacidosis Events	Diabetes Complications (retinopathy, nephropathy, neuropathy, diabetic foot)	Health-Related Quality of Life
	Baseline to 24 Wk	At 24 Wk	At 24 Wk			DTSQ Overall Mean Score at 24 Wk
Beck et al (2017) <sup>37</sup>						
N	158	158	158	158	NR	150
CGM	8.6 to 7.7	11 (14%)	40 (52%)	0		Baseline: 1.78 24 weeks: 1.61
Control	8.6 to 8.2	9 (12%)	24 (32%)	0		Baseline: 1.69 24 weeks: 1.78

Study	Reduction in HbA1c Levels (Mean Range), %	HbA1c Level <7.0%, n (%)	Relative Reduction in HbA1c Level ≥10%, n (%)	Hypoglycemic or Ketoacidosis Events	Diabetes Complications (retinopathy, nephropathy, neuropathy, diabetic foot)	Health-Related Quality of Life
TE (95% CI)	-0.3 (-0.5 to 0.0)	3% (-9% to 14%)	22% (0% to 42%)			0.22 (0.08 to 0.36)
p	.022	.88	.028			.009
Baseline to 12 Wk						
Ehrhardt et al (2011) <sup>31</sup>						
N	100	NR	NR	NR	NR	NR
CGM	8.4 to 7.4					
Control	8.2 to 7.7					
TE (95% CI)	NR					
p	.006					
Baseline to 8 months at 8 months						
Martens et al (2021) <sup>32</sup>						
NCT03566693						
N	156	156	156	175	NR	NR
CGM	9.1% to 8.0%	20 (19%)	66 (63%)	1 hypoglycemic event, 1 ketoacidosis event		
Control	9.0% to 8.4%	5 (10%)	21 (41%)	1 hypoglycemic event		
TE (95% CI)	-0.4% (-0.8% to -0.1%)	11.8 (0.6 to 24.5)	22.4 (12.0 to 32.0)			
p	.02	.04	<.001			

CGM: continuous glucose monitoring; CI: confidence interval; DDS: Diabetes Distress Scale; DTSQ: Diabetes Treatment Satisfaction; HbA1c: hemoglobin A<sub>1c</sub>; NCT: national clinical trial; NR: not reported; RCT: randomized controlled trial; TE: treatment effect.

<sup>a</sup> serious hypoglycemic event defined as requiring third-party assistance.

Tables 20 and 21 display notable limitations identified in each study.

**Table 20. Study Relevance Limitations of RCTs for Glucose Monitoring in Type 2 Diabetes**

Study; Trial	Population <sup>a</sup>	Intervention <sup>b</sup>	Comparator <sup>c</sup>	Outcomes <sup>d</sup>	Follow-Up <sup>e</sup>
					1. Follow-up not sufficient to determine effects on diabetic com
Beck et al (2017) <sup>32</sup> ; DIAMOND				1. Did not include outcomes on diabetic complications	1. Follow-up not sufficient to determine effects on diabetic complications
Ehrhardt et al (2011) <sup>31</sup>				1. Focused on HbA1c; did not include outcomes on adverse events, QOL, or diabetic complications <sup>6</sup> . No	1. Follow-up not sufficient to determine effects on diabetic

Study; Trial	Population <sup>a</sup>	Intervention <sup>b</sup>	Comparator <sup>c</sup>	Outcomes <sup>d</sup>	Follow-Up <sup>e</sup>
				justification for clinically significant difference	complications; patients reportedly followed for 52 wk but data not reported.
Martens et al (2021) <sup>39</sup>				1. Did not include outcomes on QOL or diabetic complications 24 participants could not complete clinic visits due to COVID-19 restrictions; fingerstick used instead of capillary blood draw in these cases.	1. Follow-up not sufficient to determine effects on diabetic complications

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

COVID-19: coronavirus disease 2019; HbA1c: hemoglobin A<sub>1c</sub>; QOL: quality of life; RCT: randomized controlled trial.

<sup>a</sup> Population key: 1. Intended use population unclear; 2. Clinical context is unclear; 3. Study population is unclear; 4. Study population not representative of intended use.

<sup>b</sup> Intervention key: 1. Not clearly defined; 2. Version used unclear; 3. Delivery not similar intensity as comparator; 4. Not the intervention of interest.

<sup>c</sup> Comparator key: 1. Not clearly defined; 2. Not standard or optimal; 3. Delivery not similar intensity as intervention; 4. Not delivered effectively.

<sup>d</sup> Outcomes key: 1. Key health outcomes not addressed; 2. Physiologic measures, not validated surrogates; 3. No CONSORT reporting of harms; 4. Not establish and validated measurements; 5. Clinical significant difference not prespecified; 6. Clinical significant difference not supported.

<sup>e</sup> Follow-Up key: 1. Not sufficient duration for benefit; 2. Not sufficient duration for harms.

**Table 21. Study Design and Conduct Limitations of RCTs for Glucose Monitoring in Type 2 Diabetes**

Study; Trial	Allocation <sup>a</sup>	Blinding <sup>b</sup>	Selective Reporting <sup>c</sup>	Data Completeness <sup>d</sup>	Power <sup>e</sup>	Statistical <sup>f</sup>
						3, 4. Treatment effects and CIs not calculated
Beck et al (2017) <sup>32</sup> ; DIAMOND		1. Not blinded; chance of bias in clinical management				
Ehrhardt et al (2011) <sup>31</sup>		1. Not blinded; chance of bias in clinical management	1. Registration not reported		3. No justification for difference used for power calculation	3, 4. Treatment effects and CIs not calculated
Martens et al (2021) <sup>39</sup>		1. Not blinded; chance of bias in clinical management				

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CI: confidence interval; RCT: randomized controlled trial.

<sup>a</sup> Allocation key: 1. Participants not randomly allocated; 2. Allocation not concealed; 3. Allocation concealment unclear; 4. Inadequate control for selection bias.

<sup>b</sup> Blinding key: 1. Not blinded to treatment assignment; 2. Not blinded outcome assessment; 3. Outcome assessed by treating physician.

<sup>c</sup> Selective Reporting key: 1. Not registered; 2. Evidence of selective reporting; 3. Evidence of selective publication.

<sup>d</sup> Data Completeness key: 1. High loss to follow-up or missing data; 2. Inadequate handling of missing data; 3. High number of crossovers; 4. Inadequate handling of crossovers; 5. Inappropriate exclusions; 6. Not intent to treat analysis (per protocol for noninferiority trials).

<sup>e</sup> Power key: 1. Power calculations not reported; 2. Power not calculated for primary outcome; 3. Power not based on clinically important difference.

<sup>f</sup> Statistical key: 1. Analysis is not appropriate for outcome type: (a) continuous; (b) binary; (c) time to event; 2. Analysis is not appropriate for multiple observations per patient; 3. Confidence intervals and/or p values not reported; 4. Comparative treatment effects not calculated.

### **Pregnant Women**

As discussed in the section on CGM in pregnant women, 2 RCTs have evaluated short-term glucose monitoring in pregnant women with type 1 and type 2 diabetes. Most women had type 1 diabetes in both trials. There were 25 (35%) women with type 2 diabetes in Murphy et al (2008)<sup>29</sup> and 31 (20%) with type 2 diabetes in Secher et al (2013).<sup>28</sup> Results for women with type 2 diabetes were not reported in Murphy et al (2008). Secher et al (2013) reported that 5 (17%) women with type 2 diabetes experienced 15 severe hypoglycemic events, with no difference between groups; other analyses were not stratified by diabetes type.

Only 2 RCTs used blinded CGM; in 1, there was no difference in reduction in HbA1c levels between CGM and control.

### **Section Summary: Continuous Glucose Monitoring for Use in Type 2 Diabetes**

Most RCTs of CGM in patients with type 2 trials found statistically significant benefits of CGM regarding glycemic control. However, the degree of HbA1c reduction and the difference in HbA1c reduction between groups might not be clinically significant. Moreover, additional evidence would be needed to show what levels of improvements in HbA1c levels over the short-term would be linked to meaningful improvements over the long-term in health outcomes such as diabetes-related morbidity and complications. Also, the variability in entry criteria as well as among interventions makes it difficult to identify an optimal approach to CGM use; the studies used a combination of intermittent and continuous monitoring with a review of data in real-time or at study visits only. Only the DIAMOND trial (n=158) used real-time CGM in type 2 diabetes. Selected patients were highly compliant during a run-in phase. The difference in change in HbA1c levels from baseline to 24 weeks was -0.3% favoring CGM. The difference in the proportion of patients with a relative reduction in HbA1c level by 10% or more was 22% favoring CGM. There were no differences in the proportions of patients with an HbA1c level of less than 7% at week 24. There were no events of severe hypoglycemia or diabetic ketoacidosis in either group. The treatment groups did not differ in any of the QOL measures. Two trials of CGM have enrolled pregnant women with type 2 diabetes but the total number of women with type 2 diabetes included in both trials is only 58. One study reported a difference in HbA1c levels at 36 weeks, and the proportion of infants that were large for gestational age (>90th percentile) favored CGM while the second study did not. Neither trial reported analyses stratified by diabetes type.

### **Use of Long-Term Continuous Glucose Monitoring in Individuals with Type 2 Diabetes on Multiple Daily Doses of Insulin with Significant Hypoglycemia in the Setting of Insulin Deficiency**

#### **Clinical Context and Therapy Purpose**

The purpose of long-term CGM devices is to provide a treatment option that is an alternative to or an improvement on existing therapies in individuals with type 2 diabetes (T2DM).

The question addressed in this evidence review is: Does the use of long-term CGM devices improve the net health outcome for individuals with type 2 diabetes who are on multiple daily doses of insulin with significant hypoglycemia in the setting of insulin deficiency?

The following PICO was used to select literature to inform this review.

### ***Populations***

The relevant population of interest is a subgroup of individuals with type 2 diabetes who are willing and able to use the device and have adequate medical supervision and who experience significant hypoglycemia on multiple daily doses of insulin or an insulin pump in the setting of insulin deficiency who receive long-term (continuous) glucose monitoring.

### ***Interventions***

The testing being considered is the use of long-term CGM devices to assess blood glucose levels and detect hypoglycemia as part of optimal diabetes management.

### ***Comparators***

The following practice is currently being used to measure glucose levels: SMBG (capillary blood sampling (finger stick) using blood glucose meters) and periodic measurement of HbA1c.

### ***Outcomes***

The general outcomes of interest are the frequency of and time spent in hypoglycemia, the incidence of hypoglycemic episodes, complications of hypoglycemia, and QOL. To assess short-term outcomes a minimum follow-up of 8 to 12 weeks is appropriate. To assess long-term outcomes follow-up of 6 months to 1 year would be appropriate.

### **Study Selection**

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess long-term outcomes and adverse effects, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

### **Review of Evidence**

#### **Systematic Reviews**

Meta-analytic results for long-term CGM in type 2 diabetes are summarized in Table 17. The largest and most recently published systematic review of RCTs (Ida et al [2019]<sup>30</sup>) reported a statistically significant reduction in hypoglycemic events in 285 subjects for CGM with a mean reduction of -0.35 (mean difference -0.59 to -0.10, p=.0006).

#### **Key Non-Randomized Trials**

Twelve-month open-access, follow-up results for long-term CGM in 108 individuals with type 2 diabetes treated with intensive insulin therapy are summarized in Haak et al (2017), which was included in the meta-analysis by Ida et al (2019).<sup>40</sup> Hypoglycemia was analyzed using 3 different glucose level thresholds (<70 mg/dl, <55 mg/dl, and <45 mg/dl). At all 3 glucose level thresholds, there were statistically significant reductions in time in hypoglycemia, frequency of hypoglycemic events, time in nocturnal hypoglycemia, and frequency of nocturnal hypoglycemia. Change for hypoglycemic events per day at 12 months compared to baseline was also significant: -40.8% (glucose <70 mg/dl, p<.0001); -56.5% (glucose <55 mg/dl, p<.0001); -61.7% (glucose <45 mg/dl, p=.0001).

**Section Summary: Use of Continuous Glucose Monitoring in Individuals with Type 2 Diabetes on Multiple Daily Doses of Insulin with Significant Hypoglycemia in the Setting of Insulin Deficiency**

A recently published systematic review and 12-month follow-up study using long-term CGM in patients with type 2 diabetes demonstrate that CGM can significantly reduce time in hypoglycemia and frequency of hypoglycemia events both during the day and at night. At 12-month follow-up, hypoglycemic events were reduced by 40.8% to 61.7% with a greater relative reduction in the most severe thresholds of hypoglycemia. The published evidence supports a meaningful improvement in the net health outcome.

**Continuous Glucose Monitoring Use in Pregnant Women With Gestational Diabetes  
Clinical Context and Therapy Purpose**

The purpose of long-term CGM and short-term (intermittent) glucose monitoring devices is to provide a treatment option that is an alternative to or an improvement on existing therapies in women with gestational diabetes.

The question addressed in this evidence review is: Do the use of long-term CGM and short-term (intermittent) glucose monitoring devices improve the net health outcome for women with gestational diabetes?

The following PICO was used to select literature to inform this review.

***Populations***

The relevant population of interest are women with gestational diabetes.

***Interventions***

The testing being considered are devices that provide continuous, long-term glucose levels to the patient to direct insulin regimens and intermittent (i.e., 72 hours), short-term monitoring of glucose levels used by the provider to optimize management.

***Comparators***

The following practice is currently being used to measure glucose levels: capillary blood sampling (finger stick) for blood glucose meters for self-monitoring.

***Outcomes***

The general outcomes of interest are a change in HbA1c levels, time spent in hypoglycemia, the incidence of hypoglycemic events, complications of hypoglycemia and QOL.

To assess short-term outcomes such as HbA1c levels, time spent in hypoglycemia, the incidence of hypoglycemic events and, complications of hypoglycemia, a minimum follow-up of 8 to 12 weeks is appropriate. To assess long-term outcomes such as QOL and maternal and infant outcomes, follow-up of 24 to 36 weeks would be appropriate.

**Study Selection**

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs;
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess long-term outcomes and adverse effects, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

**Review of Evidence****Randomized Controlled Trials**

One trial of glucose monitoring in women with gestational diabetes has been published. Trial characteristics, results, and limitations are shown in Tables 22 to 25. In the RCT, Wei et al (2016)

evaluated the use of CGM in 120 women with gestational diabetes at 24 to 28 weeks.<sup>41</sup> Patients were randomized to prenatal care plus CGM (n=58) or SMBG (n=62). The CGM sensors were reportedly inserted for 48 to 72 hours on weekdays; it is not clear whether the readings were available in real-time. The investigators assessed a number of endpoints and did not specify primary outcomes; a significance level of p less than .05 was used for all outcomes. The groups did not differ significantly in a change in most outcomes, including a change in maternal HbA1c levels, rates of preterm delivery before the 35th gestational week, cesarean delivery rates, proportions of large-for-gestational-age infants, or rates of neonatal hypoglycemia. Women in the CGM group gained significantly less weight than those in the SMBG group.

**Table 22. Key RCT Characteristics for CGM in Pregnant Women With Gestational Diabetes**

Study	Countries	Sites	Dates	Participants	Interventions	
					Active	Comparator
Wei et al (2016) <sup>41</sup>	China	1	2011-2012	Pregnant women with gestational diabetes diagnosed between 24 and 28 wk of gestation; mean HbA1c level, 5.8%; mean age, 30 y	CGM (48- 721 on weekdays) (n=51)	SMBG (n=55)

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; RCT: randomized controlled trial; SMBG: self-monitored blood glucose.

**Table 23. RCT Outcomes for CGM in Pregnant Women With Gestational Diabetes**

Study	Infant				Maternal	
	Large-for-Gestational Age, n (%)	Gestational Age at Delivery, wk	Severe Hypoglycemia, n (%)	Caesarean Section, n (%)	HbA1c Levels at 36 Wk of Gestation <sup>a</sup>	Severe Hypoglycemia
Wei et al (2016) <sup>41</sup>						
N	106	106	106	106		NR
CGM	18 (35)	Mean, 37.4	4 (8)	31 (60)	Mean, 5.5%	
Control	29 (53)	Mean, 37.5	7 (13)	38 (69)	Mean, 5.6%	
TE (95% CI)	NR	NR	NR	NR	NR	
p	.07	.92	.41	.37	.09	

Values are n (%) or as otherwise indicated.

CGM: continuous glucose monitoring; CI: confidence interval; HbA1c: hemoglobin A<sub>1c</sub>; NR: not reported; RCT: randomized controlled trial; TE: treatment effect.

<sup>a</sup> N inconsistently reported for HbA1c outcome.

Tables 24 and 25 display notable limitations identified in each study.

**Table 24. Study Relevance Limitations of RCTs for CGM in Pregnant Women With Gestational Diabetes**

Study	Population <sup>a</sup>	Intervention <sup>b</sup>	Comparator <sup>c</sup>	Outcomes <sup>d</sup>	Follow-Up <sup>e</sup>
Wei et al (2016) <sup>41</sup>	4. Study population had relatively low HbA1c level	4. Compliance with CGM not reported			

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; RCT: randomized controlled trial.

<sup>a</sup> Population key: 1. Intended use population unclear; 2. Clinical context is unclear; 3. Study population is unclear; 4. Study population not representative of intended use.

<sup>b</sup> Intervention key: 1. Not clearly defined; 2. Version used unclear; 3. Delivery not similar intensity as comparator; 4. Not the intervention of interest.

<sup>c</sup> Comparator key: 1. Not clearly defined; 2. Not standard or optimal; 3. Delivery not similar intensity as intervention; 4. Not delivered effectively.

<sup>d</sup> Outcomes key: 1. Key health outcomes not addressed; 2. Physiologic measures, not validated surrogates; 3. No CONSORT reporting of harms; 4. Not establish and validated measurements; 5. Clinical significant difference not prespecified; 6. Clinical significant difference not supported.

<sup>e</sup> Follow-Up key: 1. Not sufficient duration for benefit; 2. Not sufficient duration for harms.

**Table 25. Study Design and Conduct Limitations of RCTs for CGM in Pregnant Women With Gestational Diabetes**

Study	Allocation <sup>a</sup>	Blinding <sup>b</sup>	Selective Reporting <sup>c</sup>	Data Completeness <sup>d</sup>	Power <sup>e</sup>	Statistical <sup>f</sup>
Wei et al (2016) <sup>41</sup>	3. Not reported	1. Not blinded; chance of bias in clinical management	1. Registration not reported	5. Exclusions not well justified	1. No power calculations reported; primary outcome not specified	3, 4. Treatment effects and CIs not calculated

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

CGM: continuous glucose monitoring; CI: confidence interval; RCT: randomized controlled trial.

<sup>a</sup> Allocation key: 1. Participants not randomly allocated; 2. Allocation not concealed; 3. Allocation concealment unclear; 4. Inadequate control for selection bias.

<sup>b</sup> Blinding key: 1. Not blinded to treatment assignment; 2. Not blinded outcome assessment; 3. Outcome assessed by treating physician.

<sup>c</sup> Selective Reporting key: 1. Not registered; 2. Evidence of selective reporting; 3. Evidence of selective publication.

<sup>d</sup> Data Completeness key: 1. High loss to follow-up or missing data; 2. Inadequate handling of missing data; 3. High number of crossovers; 4. Inadequate handling of crossovers; 5. Inappropriate exclusions; 6. Not intent to treat analysis (per protocol for noninferiority trials).

<sup>e</sup> Power key: 1. Power calculations not reported; 2. Power not calculated for primary outcome; 3. Power not based on clinically important difference.

<sup>f</sup> Statistical key: 1. Analysis is not appropriate for outcome type: (a) continuous; (b) binary; (c) time to event; 2. Analysis is not appropriate for multiple observations per patient; 3. Confidence intervals and/or p values not reported; 4. Comparative treatment effects not calculated.

### Section Summary: Continuous Glucose Monitoring Use in Pregnant Women With Gestational Diabetes

The RCT in women with gestational diabetes was conducted in China with the intervention starting in the second or third trimester and mean baseline HbA1c level less than 6.0%. The type of CGM monitoring was unclear. Trial reporting was incomplete; however, there were no differences between groups for most reported outcomes.

### Continuous Glucose Monitoring Using Intermittently Scanned (Flash) Devices Clinical Context and Therapy Purpose

The purpose of CGM using intermittently scanned (flash) devices is to provide a treatment option that is an alternative to or an improvement on existing therapies in individuals with type 1 or type 2 diabetes.

The question addressed in this evidence review is: Does the use of intermittently scanned (flash) continuous glucose monitors improve the net health outcome in individuals with type 1 or type 2 diabetes?

The following PICO was used to select literature to inform this review.

#### Populations

The relevant population of interest are individuals with type 1 or type 2 diabetes.

#### Interventions

The testing being considered is intermittently scanned (flash) continuous glucose monitors. Flash glucose monitors measure glucose levels continuously but only display glucose values when

scanned by a reader or a smartphone. A sensor inserted on the upper arm allows high frequency monitoring of interstitial fluid for up to 14 days.

### Comparators

The comparator is capillary blood sampling (finger stick) for blood glucose meters for self-monitoring.

### Outcomes

The general outcomes of interest are a change in HbA1c levels, time spent in hypoglycemia, the incidence of hypoglycemic events, complications of hypoglycemia and QOL.

To assess short-term outcomes such as HbA1c levels, time spent in hypoglycemia, the incidence of hypoglycemic events and, complications of hypoglycemia, a minimum follow-up of 8 to 12 weeks is appropriate. To assess long-term outcomes such as QOL and maternal and infant outcomes, follow-up of 24 to 36 weeks would be appropriate.

### Study Selection

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs;
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess long-term outcomes and adverse effects, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

### Review of Evidence

#### Randomized Controlled Trials

Randomized controlled trials comparing flash glucose monitors to usual care (SMBG) are summarized in Tables 26 and 27. Two RCTs were conducted in individuals with Type 1 diabetes and 2 in individuals with Type 2 diabetes. The studies varied in their populations, primary outcome measures, and duration; further limitations are highlighted in Tables 28 and 29.

**Table 26. RCTs of Flash Glucose Monitoring in Type 1 and Type 2 Diabetes**

Study; Registration	Countries	Sites	Dates	Participants	Interventions
Bolinder et al (2016) <sup>42</sup>	Multiple European	23	2014- 2015	Adults with type 1 diabetes, well controlled (HbA1c ≤7.5%) and on current insulin regimen for at least 3 months	Flash glucose monitoring with FreeStyle Libre device n = 121 SMBG n = 120
Secher et al (2021) <sup>43</sup> NCT03682237	Denmark	5	2018- 2020	Adults with type 1 diabetes, multiple daily insulin injections and HbA1c > 7.0%	Flash glucose monitoring with FreeStyle Libre device n = 48 Usual care n = 42 Automated bolus calculation n = 41

Study; Registration	Countries	Sites	Dates	Participants	Interventions
					Automated bolus calculation plus flash monitoringn = 39
Haak et al (2017) <sup>36</sup> NCT02082184	Multiple European	26	2014-2015	Adults with type 2 diabetes treated with insulin for at least 6 months and on their current regimen for 3 months or more, HbA1c 7.5 to 12.0%.	Flash glucose monitoring with FreeStyle Libre device n = 149
					SMBG n = 75
Yaron et al (2019) <sup>44</sup>	Israel	2	2016-2017	Adults with type 2 diabetes on multiple daily insulin injections for at least 1 year.	Flash glucose monitoring with FreeStyle Libre device n = 53
					SMBG n = 48
Furler et al (2020) <sup>45</sup> NCT02082184	Australia	25	2016-2017	Adults with type 2 diabetes for at least 1 year and HbA1c at least 0.5% above their target despite being prescribed 2 non-insulin glucose-lowering drugs, insulin, or both.	Flash glucose monitoring with FreeStyle Libre Pro device (hand-held reader retained by the health professional, n = 149
					SMBG n = 150

CGM: continuous glucose monitoring; HbA1c: hemoglobin A<sub>1c</sub>; NCT: national clinical trial; RCT: randomized controlled trial; SMBG: self-monitored blood glucose.

**Table 27. Key RCT Outcomes for Glucose Monitoring in Type 2 Diabetes**

Study	Main Results - Primary Outcome Intervention vs Control	Main Results- Secondary Outcomes
Bolinder et al (2016) <sup>42</sup>	Mean time in hypoglycemia, change from baseline to 6 months: -1.39 hours/day vs -0.14 hours/day Between-group difference: -1.24 (SE 0.239; p <.0001)	5 serious adverse events in each group
Secher et al (2021) <sup>43</sup>	Time in range (3.9 to 10 mmol/l), change from baseline to 26 weeks: Between-group difference: 3.9% (95% CI -12% to -23%); p =.660	
Haak et al (2017) <sup>36</sup> NCT02082184	HbA1c change from baseline to 6 months: -3.1 (SE 0.75) mmol/L (-0.29% ± 0.07%) vs -3.4 (SE 1.04 [-0.31 ± 0.09%])	Time in hypoglycemia: <3.9 mmol/L: reduced by mean 0.47 (SE 0.13) hours/day; p =.0006

Study	Main Results - Primary Outcome Intervention vs Control	Main Results- Secondary Outcomes
Yaron et al (2019) <sup>44</sup> .	p =.8222 Treatment satisfaction (DTSQc) at 10 weeks: 2.47 (0.77) vs. 2.18 (0.83);	<3.1 mmol/L reduced by 0.22 ± 0.07 hours/day; p =.0014  The changes in HbA1c were -0.82% (9 mmol/mol) vs. -0.33% (3.6 mmol/mol) in the intervention and control group, respectively (P = 0.005);
	p =.053 Change from baseline in mean HbA1c at 12 months: 8.2% (95% CI 8.0 to 8.4) vs 8.5% (8.3 to 8.7)  Between-group difference -0.3% (95% CI, -0.5 to 0.01)  66 mmol/ml (95% CI 64 to 68) vs 69 mmol/ml (67 to 72)  Between-group difference -3.0 (95% CI, -5.0 to 0.1); p =.059	At 6 months, HbA1c was lower in the flash glucose monitoring group than in the usual care group\  Between-group difference -0.5% (95% CI, -0.8% to -0.3%); p =.0001 Clinically significant hypoglycaemia increased in both groups during the trial, but no episodes of severe hypoglycaemia or other adverse events reported in either group.

CGM: continuous glucose monitoring; CI: confidence interval; RCT: randomized controlled trial; SE: standard error.

**Table 28. Study Relevance Limitations**

Study	Population <sup>a</sup>	Intervention <sup>b</sup>	Comparator <sup>c</sup>	Outcomes <sup>d</sup>	Duration of Follow-up <sup>e</sup>
Bolinder et al (2016) <sup>42</sup> .				1. Did not include outcomes on diabetic complications	
Secher et al (2021) <sup>43</sup> .				1. Did not include outcomes on diabetic complications	
Haak et al (2017) <sup>36</sup> .				1. Did not include outcomes on diabetic complications	1. Follow-up not sufficient to determine effects on diabetic complications
Yaron et al (2019) <sup>44</sup> .				1. Did not include outcomes on diabetic complications; primary outcome was treatment satisfaction	1. Followup 10 weeks- not sufficient to determine effects on diabetic complications
Furler et al (2020) <sup>45</sup> .		5. used pro system; not currently available in the US.		1. Did not include outcomes on diabetic complications	

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

<sup>a</sup> Population key: 1. Intended use population unclear; 2. Study population is unclear; 3. Study population not representative of intended use; 4. Enrolled populations do not reflect relevant diversity; 5. Other.

<sup>b</sup> Intervention key: 1. Not clearly defined; 2. Version used unclear; 3. Delivery not similar intensity as comparator; 4. Not the intervention of interest (e.g., proposed as an adjunct but not tested as such); 5. Other.

<sup>c</sup> Comparator key: 1. Not clearly defined; 2. Not standard or optimal; 3. Delivery not similar intensity as intervention; 4. Not delivered effectively; 5. Other.

<sup>d</sup> Outcomes key: 1. Key health outcomes not addressed; 2. Physiologic measures, not validated surrogates; 3. Incomplete reporting of harms; 4. Not establish and validated measurements; 5. Clinically significant difference not prespecified; 6. Clinically significant difference not supported; 7. Other.

<sup>e</sup> Follow-Up key: 1. Not sufficient duration for benefit; 2. Not sufficient duration for harms; 3. Other.

**Table 29. Study Design and Conduct Limitations**

Study	Allocation <sup>a</sup>	Blinding <sup>b</sup>	Selective Reporting <sup>c</sup>	Data Completeness <sup>d</sup>	Power <sup>e</sup>	Statistical <sup>f</sup>
Bolinder et al (2016) <sup>42</sup>		Open label				
Secher et al (2021) <sup>43</sup>		Open label				
Haak et al (2017) <sup>36</sup>		1. Pre-randomization blinded run-in phase for both groups. Control group only blinded for last 2 weeks of study				3, 4. Treatment effects and CIs not calculated
Yaron et al (2019) <sup>44</sup>		Open label		primary outcome assessed for 86.8% of intervention group and 75% of control group		
Furler et al (2020) <sup>45</sup>		Open label				

The study limitations stated in this table are those notable in the current review; this is not a comprehensive gaps assessment.

<sup>a</sup> Allocation key: 1. Participants not randomly allocated; 2. Allocation not concealed; 3. Allocation concealment unclear; 4. Inadequate control for selection bias; 5. Other.

<sup>b</sup> Blinding key: 1. Participants or study staff not blinded; 2. Outcome assessors not blinded; 3. Outcome assessed by treating physician; 4. Other.

<sup>c</sup> Selective Reporting key: 1. Not registered; 2. Evidence of selective reporting; 3. Evidence of selective publication; 4. Other.

<sup>d</sup> Data Completeness key: 1. High loss to follow-up or missing data; 2. Inadequate handling of missing data; 3. High number of crossovers; 4. Inadequate handling of crossovers; 5. Inappropriate exclusions; 6. Not intent to treat analysis (per protocol for noninferiority trials); 7. Other.

<sup>e</sup> Power key: 1. Power calculations not reported; 2. Power not calculated for primary outcome; 3. Power not based on clinically important difference; 4. Other.

<sup>f</sup> Statistical key: 1. Analysis is not appropriate for outcome type: (a) continuous; (b) binary; (c) time to event; 2. Analysis is not appropriate for multiple observations per patient; 3. Confidence intervals and/or p values not reported; 4. Comparative treatment effects not calculated; 5. Other.

### Section Summary: Continuous Glucose Monitoring Using Intermittently Scanned (Flash) Devices

Two RCTs using the Freestyle Libre device in patients with Type 1 diabetes have been published. One RCT reported decreased time in hypoglycemia for patients with well-controlled diabetes who used flash monitors compared to SMBG (between-group difference: -1.24 hours/day

(standard error, 0.239;  $P < .0001$ ). The other study found no difference in time in range (3.9 to 10 mmol/L) for patients who used flash monitoring compared to usual care.

Three RCTs using flash glucose-sensing technology as a replacement for SMBG for the management of insulin-dependent treated type 2 diabetes have been published; 1 of these used a professional device where the healthcare professional maintained the handheld device. One RCT found no improvement on the primary outcome of A1c at 12 months or diabetes-specific distress compared to usual care and clinically significant hypoglycemia increased in both groups. However, A1c was improved at 6 months and time in target was improved at 12 months in the flash monitor group compared to usual care. A second RCT found no difference in A1c change at 6 and 12 months between flash monitor and usual care groups, but severe hypoglycemia ( $<45\text{mg/dL}$ ) was reduced for intervention participants. In a third trial, the primary outcome of patient satisfaction at 10 weeks was increased with flash monitors compared to usual care. No studies were of sufficient duration to assess diabetes complications.

The limited number of RCTs and differences across these trials in the patient populations, interventions, and outcomes, along with inconsistent results on intermediate measures, preclude drawing conclusions about the effectiveness of flash glucose monitors compared to SMBG in individuals with Type 1 or Type 2 diabetes.

## **Summary of Evidence**

### **Type 1 Diabetes**

For individuals with type 1 diabetes who are willing and able to use the device, and have adequate medical supervision, who receive long-term CGM, the evidence includes RCTs and systematic reviews. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity. Systematic reviews have generally found that at least in the short-term, long-term CGM resulted in significantly improved glycemic control for adults and children with type 1 diabetes, particularly highly compliant patients. A 2017 individual patient data analysis, pooling data from 11 RCTs, found that reductions in HbA1c levels were significantly greater with real-time CGM than with a control intervention. Two RCTs in patients who used multiple daily insulin injections and were highly compliant with CGM devices during run-in phases found that CGM was associated with a larger reduction in HbA1c levels than previous studies. One of the 2 RCTs prespecified hypoglycemia-related outcomes and reported that time spent in hypoglycemia was significantly less in the CGM group. One RCT in pregnant women with type 1 diabetes, which compared real-time CGM with self-monitoring of blood glucose, has also reported a difference in change in HbA1c levels, an increased percentage of time in the recommended glucose control target range, a smaller proportion of infants who were large for gestational age, a smaller proportion of infants who had neonatal intensive care admissions lasting more than 24 hours, a smaller proportion of infants who had neonatal hypoglycemia requiring treatment, and reduced total hospital length of stay all favoring CGM. The evidence is sufficient to determine that the technology results in an improvement in the net health outcome.

For individuals with type 1 diabetes who receive short-term continuous glucose monitoring, the evidence includes RCTs and systematic reviews. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity as well as intermediate outcomes related to measures of glucose control such as frequency and time in hypoglycemia and hyperglycemia. The evidence for short-term monitoring of glycemic control is mixed, and there was no consistency in HbA1c levels. Some trials have reported improvements in glucose control for the intermittent short-term monitoring group but limitations in this body of evidence preclude conclusions. The definitions of control with short-term CGM use, duration of use and the specific monitoring protocols varied. In some studies, short-term monitoring was part of a larger strategy aimed at optimizing glucose control, and the impact of monitoring cannot be separated from the impact of other interventions. Studies have not shown an advantage for intermittent glucose monitoring in reducing severe hypoglycemia events but the number of events reported is generally small and effect estimates imprecise. The limited duration of use may preclude an assessment of any therapeutic effect. Two RCTs of short-term CGM use for monitoring in

pregnancy included women with both type 1 and 2 diabetes, with most having type 1 diabetes. One trial reported a difference in HbA1c levels at 36 weeks; the proportion of infants that were large for gestational age (>90th percentile) favored CGM while the second trial did not. The differences in the proportions of infants born via cesarean section, gestational age at delivery, and infants with severe hypoglycemia were not statistically significant in either study. Limitations of the published evidence preclude determining the effects of the technology on net health outcome. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

For individuals with type 1 diabetes who receive continuous glucose monitoring using an intermittently scanned (flash) device, the evidence includes RCTs. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity as well as intermediate outcomes related to measures of glucose control such as frequency and time in hypoglycemia and hyperglycemia. Two RCTs using the Freestyle Libre device have been published. One RCT reported decreased time in hypoglycemia for patients with well-controlled diabetes who used flash monitors compared to SMBG (between-group difference: -1.24 hours/day [standard error, 0.239; P <.0001]). The other study found no difference in time in range (3.9 to 10 mmol/L) for patients who used flash monitoring compared to usual care. The limited number of RCTs and differences in their populations, primary outcomes, and comparators preclude drawing conclusions about the use of flash monitoring in patients with Type 1 diabetes. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

### **Type 2 Diabetes**

For individuals with type 2 diabetes who receive long-term CGM, the evidence includes RCTs. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity. Most RCTs of CGM in patients with type 2 diabetes found statistically significant benefits of CGM regarding glycemic control. However, the degree of HbA1c reduction and the difference in HbA1c reduction between groups might not be clinically significant. Moreover, additional evidence would be needed to show what levels of improvements in HbA1c levels over the short-term would be linked to meaningful improvements over the long-term in health outcomes such as diabetes-related morbidity and complications. Also, the variability in entry criteria as well as among interventions makes it difficult to identify an optimal approach to CGM use; the studies used a combination of intermittent and continuous monitoring with a review of data in real-time or at study visits only. Only the DIAMOND RCT (N =158) has used real-time CGM in type 2 diabetes. Selected patients were highly compliant during a run-in phase. The difference in change in HbA1c levels from baseline to 24 weeks was -0.3% favoring CGM. The difference in the proportion of patients with a relative reduction in HbA1c level by 10% or more was 22% favoring CGM. There were no differences in the proportions of patients with an HbA1c level of less than 7% at week 24. There were no events of severe hypoglycemia or diabetic ketoacidosis in either group. The treatment groups did not differ in any of the QOL measures. RCTs using flash glucose-sensing technology as a replacement for SMBG for the management of insulin-dependent treated type 2 diabetes found no difference in HbA1c change at 6 and 12 months between groups. However, time in severe hypoglycemia (<45mg/dL) was reduced for intervention participants. Two trials of CGM have enrolled pregnant women with type 2 diabetes, but the total number of women with type 2 diabetes included in both trials is only 58. One study reported a difference in HbA1c levels at 36 weeks, and the proportion of infants that were large for gestational age (>90th percentile) favored CGM while the second study did not. Neither trial reported analyses stratified by diabetes type. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

For individuals with type 2 diabetes who are willing and able to use the device and have adequate medical supervision and who experience significant hypoglycemia on multiple daily doses of insulin or an insulin pump in the setting of insulin deficiency who receive long-term (continuous) glucose monitoring, the evidence includes a systematic review and non-randomized study with 12-month follow-up. Relevant outcomes are the frequency of and time

spent in hypoglycemia, the incidence of hypoglycemic episodes, complications of hypoglycemia, and QOL. The available studies demonstrate that CGM can significantly reduce time in hypoglycemia and frequency of hypoglycemia events both during the day and at night. At 12-month follow-up, hypoglycemic events were reduced by 40.8% to 61.7% with a greater relative reduction in the most severe thresholds of hypoglycemia. The published evidence supports a meaningful improvement in the net health outcome. Evidence reported through clinical input provides additional clinical context and based on both the published evidence and clinical input the following patient selection criteria are associated with a clinically meaningful improvement in net health outcome and are consistent with generally accepted medical practice: selected patients with type 2 diabetes who are (1) willing and able to use the CGM device and have adequate medical supervision and (2) experiencing significant hypoglycemia on multiple daily doses of insulin or an insulin pump in the setting of insulin deficiency. The evidence is sufficient to determine that the technology results in an improvement in the net health outcome.

For individuals with type 2 diabetes who receive short-term CGM monitoring, the evidence includes RCTs and systematic reviews. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity. Systematic reviews of 3 to 4 RCTs have found statistically significant benefits from CGM regarding glycemic control. However, the degree of HbA1c reduction and the difference in HbA1c reductions between groups may not be clinically significant. Also, the limited number of RCTs and variability among interventions make it difficult to identify an optimal approach to CGM or a subgroup of type 2 diabetes patients who might benefit. Moreover, studies of CGM in patients with type 2 diabetes have generally not addressed the clinically important issues of severe hypoglycemia and diabetic complications. Very few pregnant women with type 2 diabetes have been included in RCTs. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

For individuals with type 2 diabetes who receive short-term continuous glucose monitoring, the evidence includes RCTs and systematic reviews. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity as well as intermediate outcomes related to measures of glucose control such as frequency and time in hypoglycemia and hyperglycemia. The evidence for short-term monitoring of glycemic control is mixed, and there was no consistency in HbA1c levels. Some trials have reported improvements in glucose control for the intermittent short-term monitoring group but limitations in this body of evidence preclude conclusions. The definitions of control with short-term CGM use, duration of use and the specific monitoring protocols varied. In some studies, short-term monitoring was part of a larger strategy aimed at optimizing glucose control, and the impact of monitoring cannot be separated from the impact of other interventions. Studies have not shown an advantage for intermittent glucose monitoring in reducing severe hypoglycemia events but the number of events reported is generally small and effect estimates imprecise. The limited duration of use may preclude an assessment of any therapeutic effect. Two RCTs of short-term CGM use for monitoring in pregnancy included women with both type 1 and 2 diabetes, with most having type 1 diabetes. One trial reported a difference in HbA1c levels at 36 weeks; the proportion of infants that were large for gestational age (>90th percentile) favored CGM while the second trial did not. The differences in the proportions of infants born via cesarean section, gestational age at delivery, and infants with severe hypoglycemia were not statistically significant in either study. Limitations of the published evidence preclude determining the effects of the technology on net health outcome. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

For individuals with type 2 diabetes who receive continuous glucose monitoring using an intermittently scanned (flash) device, the evidence includes RCTs. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity as well as intermediate outcomes related to measures of glucose control such as frequency and time in hypoglycemia and hyperglycemia. Three RCTs using flash glucose-sensing technology as a replacement for

SMBG for the management of insulin-dependent treated type 2 diabetes have been published; 1 of these used a professional device where the healthcare professional maintained the handheld device. One RCT found no improvement on the primary outcome of A1c at 12 months or diabetes-specific distress compared to usual care and clinically significant hypoglycemia increased in both groups. However, A1c was improved at 6 months and time in target was improved at 12 months in the flash monitor group compared to usual care. A second RCT found no difference in A1c change at 6 and 12 months between flash monitor and usual care groups, but severe hypoglycemia (<45mg/dL) was reduced for intervention participants. In a third trial, the primary outcome of patient satisfaction at 10 weeks was increased with flash monitors compared to usual care. No studies were of sufficient duration to assess diabetes complications. Differences across these trials in the patient populations, interventions, and outcomes, along with inconsistent results on intermediate measures, preclude drawing conclusions about the effectiveness of flash glucose monitors compared to SMBG. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

### **Gestational Diabetes**

For individuals who are pregnant with gestational diabetes who receive long-term CGM or short-term (intermittent) glucose monitoring, the evidence includes an RCT. Relevant outcomes are symptoms, morbid events, QOL, and treatment-related morbidity. In the RCT, the type of glucose monitoring was unclear. Trial reporting was incomplete; however, there was no difference between the groups for most reported outcomes. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

### **Supplemental Information**

The purpose of the following information is to provide reference material. Inclusion does not imply endorsement or alignment with the evidence review conclusions.

### **Clinical Input From Physician Specialty Societies and Academic Medical Centers**

While the various physician specialty societies and academic medical centers may collaborate with and make recommendations during this process, through the provision of appropriate reviewers, input received does not represent an endorsement or position statement by the physician specialty societies or academic medical centers, unless otherwise noted.

### **2019 Input**

Clinical input was sought to help determine whether the use of continuous or intermittent monitoring of glucose in the interstitial fluid would provide a clinically meaningful improvement in net health outcome and whether the use is consistent with generally accepted medical practice. In response to requests, clinical input was received from 3 respondents, including 3 physician-level responses identified through 1 specialty society, including 2 physicians with academic medical center affiliations.

### **Type 1 Diabetes**

For individuals who have type 1 diabetes who receive short-term glucose monitoring, clinical input supports that this use provides a clinically meaningful improvement in net health outcome and is consistent with generally accepted medical practice when used in specific situations such as poor control of Type 1 diabetes despite the use of best practices and to help determine basal insulin levels prior to insulin pump initiation.

### **Type 2 Diabetes**

For individuals who have type 2 diabetes who do not require insulin who receive long-term continuous glucose monitoring (CGM), clinical input does not support a clinically meaningful improvement in net health outcome and does not indicate this use is consistent with generally accepted medical practice.

For individuals with type 2 diabetes who are willing and able to use the device and have adequate medical supervision and who experience significant hypoglycemia on multiple daily

doses of insulin or an insulin pump in the setting of insulin deficiency who receive long-term continuous glucose monitoring, clinical input supports that this use provides a clinically meaningful improvement in net health outcome and is consistent with generally accepted medical practice.

For individuals with type 2 diabetes who require multiple daily doses of insulin who receive short-term CGM, clinical input supports that this use provides a clinically meaningful improvement in net health outcome and is consistent with generally accepted medical practice when used in specific situations such as poor control of diabetes despite use of best practices and to help determine basal insulin levels prior to insulin pump initiation.

Further details from clinical input are included in the Appendix.

## **2008**

In response to requests from Blue Cross Blue Shield Association, input was received from 1 physician specialty society and 4 academic medical centers in 2008. Input concurred that continuous glucose monitoring, particularly intermittent glucose monitoring, was helpful in a subset of patients with diabetes. Reviewers commented that this monitoring can improve diabetes care by reducing glucose levels (and improving hemoglobin A<sub>1c</sub> levels) and/or by reducing episodes of hypoglycemia. Reviewers argued that there is persuasive data from case reports to demonstrate the positive impact of intermittent glucose monitoring.

### **Practice Guidelines and Position Statements**

Guidelines or position statements will be considered for inclusion in 'Supplemental Information' if they were issued by, or jointly by, a US professional society, an international society with US representation, or National Institute for Health and Care Excellence (NICE). Priority will be given to guidelines that are informed by a systematic review, include strength of evidence ratings, and include a description of management of conflict of interest.

### **American Association of Clinical Endocrinologists and the American College of Endocrinology**

In 2020, the American Association of Clinical Endocrinologists (AACE) and the American College of Endocrinology (ACE) 2015 Clinical Practice Guidelines for Developing a Diabetes Mellitus Comprehensive Care Plan was supplemented by an AACE/ACE Consensus Statement on Comprehensive Type 2 Diabetes Management. It is recommended that therapy be evaluated regularly including the results of hemoglobin A<sub>1c</sub>, self-monitoring of blood glucose (SMBG) records (fasting and postprandial) or continuous glucose monitoring tracings. The statement supports consideration of the use of personal CGM devices for those patients who are on intensive insulin therapy (3 to 4 injections/day or on an insulin pump), for those with a history of hypoglycemia unawareness, or those with recurrent hypoglycemia. Regarding the duration of use the statement reads; "While these devices could be used intermittently in those who appear stable on their therapy, most patients will need to use this technology on a continual basis". [46](#).

### **National Institute for Health and Care Excellence**

The National Institute for Health and Care Excellence (2016) updated its guidance on the diagnosis and management of type 1 diabetes in adults.[47](#) The guidance stated that real-time CGM should not be offered "routinely to adults with type 1 diabetes" but that it can be considered in the following:

"...adults with type 1 diabetes who are willing to commit to using it at least 70% of the time and to calibrate it as needed, and who have any of the following despite optimised use of insulin therapy and conventional blood glucose monitoring:

- More than 1 episode a year of severe hypoglycaemia with no obviously preventable precipitating cause.
- Complete loss of awareness of hypoglycaemia.
- Frequent (more than 2 episodes a week) asymptomatic hypoglycaemia that is causing problems with daily activities.
- Extreme fear of hypoglycaemia.

- Hyperglycaemia (HbA1c [hemoglobin A<sub>1c</sub>] level of 75 mmol/mol [9%] or higher) that persists despite testing at least 10 times a day. Continue real-time continuous glucose monitoring only if HbA1c can be sustained at or below 53 mmol/mol (7%) and/or there has been a fall in HbA1c of 27 mmol/mol (2.5%) or more."

### American Diabetes Association

The American Diabetes Association (2021) "Standards of Medical Care in Diabetes<sup>48</sup>" included the following statement in the chapter on glycemic targets:

*"Continuous glucose monitoring (CGM) also has an important role in assessing the effectiveness and safety of treatment in many patients with type 1 diabetes, and limited data suggest it may also be helpful in selected patients with type 2 diabetes, such as those on intensive insulin regimens."*

*The Standards also state that the technology has evolved rapidly in both accuracy and affordability and that data provided by CGM "will allow the provider to determine time in range (TIR) and to assess hypoglycemia, hyperglycemia, and glycemic variability", noting that there is a strong correlation between TIR and an A1C.*

Recommendations (**level of evidence**) related to CGM devices include the following:

- "When prescribing continuous glucose monitoring (CGM) devices, robust diabetes education, training, and support are required for optimal CGM device implementation and ongoing use. People using CGM devices need to have the ability to perform self-monitoring of blood glucose in order to calibrate their monitor and/or verify readings if discordant from their symptoms. **(B)**.
- When used properly, real-time continuous glucose monitors in conjunction with multiple daily injections and continuous subcutaneous insulin infusion **(A)** and other forms of insulin therapy **(C)** are a useful tool to lower and/or maintain A1C levels and/or reduce hypoglycemia in adults and youth with diabetes.
- When used properly, intermittently scanned continuous glucose monitors in conjunction with multiple daily injections and continuous subcutaneous insulin infusion **(B)** and other forms of insulin therapy **(C)** can be useful and may lower A1C levels and/or reduce hypoglycemia in adults and youth with diabetes to replace self-monitoring of blood glucose.
- In patients on multiple daily injections and continuous subcutaneous insulin infusion, real-time continuous glucose monitoring (CGM) devices should be used as close to daily as possible for maximal benefit. **(A)** Intermittently scanned CGM devices should be scanned frequently, at a minimum once every 8 h.
- When used as an adjunct to pre- and postprandial self-monitoring of blood glucose, continuous glucose monitoring can help to achieve A1C targets in diabetes and pregnancy. **(B)**
- Use of professional continuous glucose monitoring (CGM) and/or intermittent real-time or intermittently scanned CGM can be helpful in identifying and correcting patterns of hyper- and hypoglycemia and improving A1C levels in people with diabetes on noninsulin as well as basal insulin regimens. **(C)**
- Skin reactions, either due to irritation or allergy, should be assessed and addressed to aid in successful use of devices. **(E)**
- People who have been using continuous glucose monitors should have continued access across third-party payers. **(E)**"

### Endocrine Society

The Endocrine Society (2016) published clinical practice guidelines that included the following recommendations on CGM<sup>49</sup>:

6. "Real-time continuous glucose monitors in adult outpatients
  - 6.1 We recommend real-time continuous glucose monitoring (RT-CGM) devices for adult patients with T1DM [type 1 diabetes mellitus] who have A1C levels above target and who are willing and able to use these devices on a nearly daily basis.

6.2 We recommend RT-CGM devices for adult patients with well-controlled T1DM who are willing and able to use these devices on a nearly daily basis.

Use of continuous glucose monitoring in adults with type 2 diabetes mellitus [T2DM]

6.3 We suggest short-term, intermittent RT-CGM use in adult patients with T2DM (not on prandial insulin) who have A1C levels  $\geq 7\%$  and are willing and able to use the device."

### International Consensus on Time in Range

In 2019, consensus recommendations on clinical targets for CGM data interpretation were published and endorsed by the American Diabetes Association, American Association of Diabetes Educators, European Association for the Study of Diabetes, Foundation of European Nurses in Diabetes, International Society for Pediatric and Adolescent Diabetes, JDRF, and Pediatric Endocrine Society.<sup>50</sup>

### U.S. Preventive Services Task Force Recommendations

Not applicable.

### Medicare National Coverage

In January 2017, the Centers for Medicare & Medicaid Services (CMS) ruled that CGM devices (therapeutic CGMs) approved by the U.S. Food and Drug Administration (FDA) that can be used to make treatment decisions are considered durable medical equipment.<sup>51</sup> A CGM is considered a therapeutic CGM if it is approved by the FDA for use in place of a blood glucose monitor for making diabetes treatment decisions such as changes in diet and insulin dosage. Initially, CMS did not consider the smartphone application as a DME component and did allow payment for that part of the CGM system. Subsequently, in June 2018, CMS made an announcement that Medicare's published coverage policy for CGMs will be modified to support the use of CGMs in conjunction with a smartphone, including the important data sharing function they provide for patients and their families.<sup>52</sup> Currently marketed therapeutic CGM systems are included in Table 1.

In 2020, Medicare assigned relative value units to the insertion, removal and removal/reinsertion codes uses for provision of the implantable glucose sensor device.

### Ongoing and Unpublished Clinical Trials

Some currently ongoing and unpublished trials that might influence this review are listed in Table 30.

**Table 30. Summary of Key Trials**

NCT No.	Trial Name	Planned Enrollment	Completion Date
<i>Ongoing</i>			
NCT03981328	The Effectiveness of Real Time Continuous Glucose Monitoring to Improve Glycemic Control and Pregnancy Outcome in Patients With Gestational Diabetes Mellitus	372	Oct 2021 (last update June 2019, not yet recruiting)
NCT03908125 <sup>a</sup>	A Post- Approval Study to Evaluate the Long-term Safety and Effectiveness of the Eversense® Continuous Glucose Monitoring (CGM) System	400	Mar 2023
NCT04269655 <sup>a</sup>	Scripps Digital Diabetes: Cloud-Based Continuous Glucose Monitoring (CB CGM)	300	Feb 2024
NCT04535830	The Effectiveness of Flash Glucose Monitoring System on Glycemic Control in Patients With New-onset Type 2 Diabetes#A Randomized Controlled Trial	200	Sep 2021
NCT03522870	Effects of Novel Flash Glucose Monitoring System on Glycemic Control in Adult Patients With Type 1 Diabetes Mellitus	104	Dec 2021
<i>Unpublished</i>			

NCT No.	Trial Name	Planned Enrollment	Completion Date
NCT03808376 <sup>a</sup>	PROMISE Study: A Prospective, Multicenter Evaluation of Accuracy and Safety of an Implantable Continuous Glucose Sensor Lasting up to 180 Days	181	May 2020 (actual)
NCT03445065 <sup>a</sup>	Benefits of a Long Term Implantable Continuous Glucose Monitoring System for Adults With Diabetes - France Randomized Clinical Trial	239	Aug 2020

NCT: national clinical trial.

<sup>a</sup> Denotes industry-sponsored or cosponsored trial.

## Appendix 1

### Clinical Input Respondents

Clinical input was provided by the following specialty societies and physician members identified by a specialty society or clinical health system:

- Chaitanya Mamillapalli, MD, MRCP, FAPCR, Endocrinology, Springfield Clinic, identified by American Association of Clinical Endocrinologists (AACE)\*\*
- Vijay Shivaswamy, MBBS, Endocrinology, University of Nebraska Medical Center and Omaha Veterans Administration, identified by AACE
- Janet B. McGill, MD, Endocrinology, Washington University School of Medicine, identified by AACE\*\*

\* Indicates that no response was provided regarding conflicts of interest related to the topic where clinical input is being sought.

\*\* Indicates that conflicts of interest related to the topic where clinical input is being sought were identified by this respondent (see Appendix).

Clinical input provided by the specialty society at an aggregate level is attributed to the specialty society. Clinical input provided by a physician member designated by a specialty society or health system is attributed to the individual physician and is not a statement from the specialty society or health system. Specialty society and physician respondents participating in the Evidence Street® clinical input process provide a review, input, and feedback on topics being evaluated by Evidence Street. However, participation in the clinical input process by a specialty society and/or physician member designated by a specialty society or health system does not imply an endorsement or explicit agreement with the Evidence Opinion published by BCBSA or any Blue Plan.

### CI - Objective

Clinical input was sought to help determine whether the use of a continuous or intermittent monitoring of glucose in interstitial fluid in individuals with diabetes would provide a clinically meaningful improvement in net health outcome and whether the use is consistent with generally accepted medical practice.

The following PICO applies to this request.

Populations	Interventions	Comparators	Outcomes
Individuals: <ul style="list-style-type: none"> <li>• With type 1 diabetes</li> </ul>	Interventions of interest are: <ul style="list-style-type: none"> <li>• Short-term (intermittent) glucose monitoring</li> </ul>	Comparators of interest are: <ul style="list-style-type: none"> <li>• Self-monitoring of blood glucose</li> </ul>	Relevant outcomes include: <ul style="list-style-type: none"> <li>• Symptoms</li> <li>• Morbid events</li> <li>• Quality of life</li> <li>• Treatment-related morbidity</li> </ul>



## Detailed Responses

## Respondent Profile

Physician					
#	Name	Degree	Institutional Affiliation	Clinical Specialty	Board Certification and Fellowship Training
<b>Identified by American Association of Clinical Endocrinologists</b>					
1	Chaitanya Mamillapalli	MD, MRCP, FAPCR	Springfield Clinic	Endocrinology	Endocrinology and Internal Medicine, Fellowship in Endocrinology
2	Vijay Shivaswamy	MBBS	University of Nebraska Medical Center and Omaha Veterans Administration	Endocrinology	Endocrinology
3	Janet B. McGill	MD	Washington University School of Medicine	Endocrinology	Endocrinology, Diabetes and Metabolism

## Respondent Conflict of Interest Disclosure

#	1) Research support related to the topic where clinical input is being sought		2) Positions, paid or unpaid, related to the topic where clinical input is being sought		3) Reportable, more than \$1,000, healthcare-related assets or sources of income for myself, my spouse, or my dependent children related to the topic where clinical input is being sought		4) Reportable, more than \$350, gifts or travel reimbursements for myself, my spouse, or my dependent children related to the topic where clinical input is being sought	
	YES/NO	Explanation	YES/NO	Explanation	YES/NO	Explanation	YES/NO	Explanation
1	Yes	Our team has research grant approved from Abbott diabetes care for the project "Impact of Joint Utilization of the FreeStyle Libre Flash Glucose Monitoring System and Flairz Health App on Glycemic Control in Patients with Type 2 Diabetes". This study will evaluate the concurrent use of the Abbott FreeStyle Libre Flash personal CGM and Flairz diabetes app on glycemic control for patients with type 2 diabetes mellitus requiring basal and prandial insulin for treatment	No		No		No	
2	No		No		No		No	
3	Yes	Current studies include WISDM (Wireless Innovation for Seniors	No		Yes	Intermittent speaker for Dexcom;	Yes	Travel to speaker training

#	1) Research support related to the topic where clinical input is being sought		2) Positions, paid or unpaid, related to the topic where clinical input is being sought		3) Reportable, more than \$1,000, healthcare-related assets or sources of income for myself, my spouse, or my dependent children related to the topic where clinical input is being sought		4) Reportable, more than \$350, gifts or travel reimbursements for myself, my spouse, or my dependent children related to the topic where clinical input is being sought	
	YES/NO	Explanation	YES/NO	Explanation	YES/NO	Explanation	YES/NO	Explanation
		with Diabetes Mellitus) and MOBILE (CGM in T2DM using basal insulin). Also Medtronic 670G study. Past studies include DIAMOND.				grant funding from Dexcom and Medtronic		meeting and to investigator meeting for WISDM this year.

Individual physician respondents answered at individual level. Specialty Society respondents provided aggregate information that may be relevant to the group of clinicians who provided input to the Society-level response. NR = not reported

**Questions and Responses**

- We are seeking your opinion on whether using the interventions for the below indications provide a clinically meaningful improvement in net health outcome. Please respond based on the evidence and your clinical experience. Please address these points in your response:
  - Relevant clinical scenarios (e.g., a chain of evidence) where the technology is expected to provide a clinically meaningful improvement in net health outcome;
  - Specific outcomes that are clinically meaningful;
  - Any relevant patient inclusion/exclusion criteria or clinical context important to consider in identifying individuals for this indication; and
  - Supporting evidence from the authoritative scientific literature (please include PMID).

#	Indications	Rationale-CI
1	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes	<p>Short term CGM trial for 1-2 weeks is a common practice among endocrinologists in the following situations:</p> <ul style="list-style-type: none"> <li>In patients who has frequent hypoglycemia episodes</li> <li>In patients with discrepancy between A1c and Self monitored blood glucose data</li> <li>Uncontrolled diabetes</li> <li>To monitor the trends of hyperglycemia and hypoglycemia to guide insulin dosing decisions</li> <li>To screen for occult hypoglycemia episodes in patients with hypoglycemic awareness</li> <li>Preoperative optimization of glucose control before surgery</li> <li>Post-operative optimization of glucose control after surgery</li> <li>Basal Insulin titration with new insulin pump starts</li> <li>Patients who have recurrent DKAs</li> <li>During pregnancy and to optimize glucose levels in the preconception period</li> </ul> <p>I respectfully disagree with the statement "Also, many of the trials of intermittent monitoring have included additional interventions to optimize glucose control (e.g., education, lifestyle modifications)."</p> <p>Outcome of CGMS studies should be studied in combination with interventions, as just measuring glucose data without making an intervention would not be successful.</p>
2	Use of short-term	No comments. Agree with analysis

#	Indications	Rationale-CI
	(intermittent) glucose monitoring for individuals with type 1 diabetes	Evidence is not only limited regarding short-term intermittent glucose monitoring, but the GOLD study clearly showed that when patients who have used long-term monitoring are returned to fingerstick glucose testing, their HbA1c increases back to the starting level and gains have been lost. While there are many factors driving high HbA1c in persons with T1DM, fear of hypoglycemia is the most commonly stated reason by my patients. Perkins et al recently reported that the strongest risk factor for macroalbuminuria or reduced eGFR in persons with T1DM in the DCCT/EDIC cohort is lifetime mean HbA1c, HR 1.952 per 1% higher level (95% CI 1.714 -2.223). <ul style="list-style-type: none"> <li>Perkins BA, Bebu I, de Boer IH, et al. Risk Factors for Kidney Disease in Type 1 Diabetes. Diabetes Care. 2019 42(5):883-890. PMID 30833370</li> </ul>
3	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes	Reducing glycemic burden safely requires the use of continuous glucose monitoring in the vast majority of patients. Baseline CGM data from the WISDM study (unpublished) uncovered an alarming rate of hypoglycemia in persons over the age of 60 with T1DM (unpublished data). The WISDM study will be presented at ADA and published thereafter. Long-term use of CGM in persons who are insulin deficient and require life-sustaining exogenous insulin is necessary for both safety to help achieve glycemic targets to prevent long-term complication. Short-term use of CGM is of modest benefit in persons with T1DM, may be helpful for adjusting insulin doses, but not for prevention of hypoglycemia or reduction of complications. <ol style="list-style-type: none"> <li>Review states "Four of the 6 RCTs of CGM in type 2 diabetes reported a statistically significant larger decrease in HbA1c levels with CGM than with control. No trials reported on follow-up beyond 6 months. Thus, the effect of CGM on outcomes related to diabetic complications is unknown." I respectfully disagree with the above statement.                             <ul style="list-style-type: none"> <li>It is important to note that the patients enrolled in this clinical trials have advanced diabetes with the duration of type 2 diabetes in the 8-21 years range. It is a challenge to improve glycemic control in this group, despite which CGM use in this population was associated with a meaningful improvement in A1c.</li> <li>The short duration of studies did not allow for the evaluation of long-term outcomes, but one can extrapolate from the diabetes outcome clinical trials, that the A1c improvements will help with the prevention of diabetes complications.</li> </ul> </li> <li>Additional efficacy studies which have demonstrated meaningful improvements in type 2 diabetes patients                             <ul style="list-style-type: none"> <li>Anjana RM, Kesavadev J, Neeta D, et al. A Multicenter Real-Life Study on the Effect of Flash Glucose Monitoring on Glycemic Control in Patients with Type 1 and Type 2 Diabetes. Diabetes Technol Ther. 2017;19(9):533-540. PMID: 28930495</li> </ul> </li> </ol>
1	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes	FREESTYLE LIBREPRO™ FLASH GLUCOSE MONITORING (FCGM) IN BOTH TYPE1 AND TYPE 25072 patients with diabetes (Both Type 1 and Type 2) who had an A1c ≥7%. Overall, the magnitude of reduction in the intervention group was 1% compared to 0.7% in the control group P < 0.001). The overall reduction in A1c among cases was higher in T2D (9.2% to 8.3%) compared with T1D (9.6% to 9.4%); <ul style="list-style-type: none"> <li>Yaron M, Roitman E, Aharon-Hananel G, et al. Effect of Flash Glucose Monitoring Technology on Glycemic Control and Treatment Satisfaction in Patients With Type 2 Diabetes. Diabetes Care. 2019; pii:dc180166. [Epub ahead of print] PMID 31036546</li> </ul> <p>FLASH GLUCOSE MONITORING (FGM) SYSTEM IN PATIENTS WITH TYPE 2 DIABETICS.</p> <p>The changes in HbA1c were -0.82% (9 mmol/mol) vs -0.33% (3.6 mmol/mol) in the</p>

#	Indications	Rationale-CI
		<p>intervention and control group, respectively (P = 0.005); 68.6% of the patients in the intervention group had their HbA1c reduced by <math>\geq 0.5\%</math> (5.5 mmol/mol) compared with 30.2% in the control group (P &lt; 0.001), 39.2% had their HbA1c reduced by <math>\geq 1.0\%</math> (10.9 mmol/mol) vs 18.6% in the control group (P = 0.0023) without an increased frequency of hypoglycemia. Satisfaction using the FGM system was high. The intervention group found treatment was more flexible (P = 0.019) and would recommend it to their counterparts (P = 0.023).</p>
		<p>3. CGM Improves patient adherence to the ADA recommended guidelines for glucose measurements</p> <ul style="list-style-type: none"> <li>o American Diabetes Association. Standards of medical care in diabetes – 2018. Diabetes Care. 2018. p. 41.</li> </ul> <p>In patients on intensive insulin treatment, ADA recommends checking glucose levels, but the frequency of blood glucose checking is suboptimal in the 2.6 range. The underlying cause for this is multifactorial 1. pain, 2. invasiveness 3. inconvenience, and 4. social stigma of finger stick glucose checks</p> <p>CGM addresses these drawbacks and will help with improving the frequency of glucose checks.</p> <ul style="list-style-type: none"> <li>o Dunn TC, Xu Y, Hayter G et al. Real-world flash glucose monitoring patterns and associations between self-monitoring frequency and glycaemic measures: A European analysis of over 60 million glucose tests. Diabetes Res Clin Pract. 2018 Mar;137:37-46. PMID 29278709</li> </ul> <p>Patient performed average of 16.3 scans/day in a study involving Flash Glucose CGM. Higher A1c reduction was achieved with higher frequency of scans.</p>
		<p>4. Use of alternative metrics other than A1C</p> <p>Recent studies have demonstrated glycemic variability as an independent risk factor for diabetes complications.</p> <ul style="list-style-type: none"> <li>o Monnier L, Mas E, Ginet C, et al. Activation of oxidative stress by acute glucose fluctuations compared with sustained chronic hyperglycemia in patients with type 2 diabetes. JAMA 2006;295:1681–1687. PMID: 16609090</li> <li>o Ceriello A, Esposito K, Piconi L, et al. Oscillating glucose is more deleterious to endothelial function and oxidative stress than mean glucose in normal and type 2 diabetic patients. Diabetes 2008;57:1349–1354. PMID: 18299315</li> <li>o Chang CM, Hsieh CJ, Huang JC, et al. Acute and chronic fluctuations in blood glucose levels can increase oxidative stress in type 2 diabetes mellitus. Acta Diabetol 2012;49(Suppl. 1): S171–S177. PMID: 22547264</li> </ul> <p>Use of CGM was associated with decrease in glycemic variability</p> <ul style="list-style-type: none"> <li>o Haak T, Hanaire H, Ajjan R et al. Flash Glucose-Sensing Technology as a Replacement for Blood Glucose Monitoring for the Management of Insulin-Treated Type 2 Diabetes: a Multicenter, Open-Label Randomized Controlled Trial. Diabetes Ther (2017 Feb) 8(1):55-73. PMID: 28000140</li> </ul>
		<p>5. Hypoglycemia outcomes.</p> <p>In 2009 Hypoglycemia hospitalization costs in the USA was over 4.7 billion dollars in the US. (248,422 T2 diabetes hospitalizations and 20,839 Type 1 diabetes)</p> <ul style="list-style-type: none"> <li>o Singh G, Mithal A, Mannalithara A, et al. Hospitalisations due to severe hypoglycaemia in patients with type 2 diabetes: a US national perspective. Abstract accessible at URL: <a href="https://distribute.m-anage.com/check.pic?path=events%5C176%5Cabstrac%5C25381%5C343775_628.pdf">https://distribute.m-anage.com/check.pic?path=events%5C176%5Cabstrac%5C25381%5C343775_628.pdf</a></li> </ul> <p>Use of CGM has been associated with reduction of hypoglycemia episodes.</p> <p>In the REPLACE trial Use of Flash CGM has been associated with reduction in</p>

#	Indications	Rationale-CI
		hypoglycemia episodes.  Time in hypoglycemia: <ul style="list-style-type: none"> <li>• &lt;70 mg/dL reduced by <math>0.47 \pm 0.13</math> h/day [mean <math>\pm</math> SE (<math>p = 0.0006</math>)] 43% reduction</li> <li>• &lt;55 mg/dL reduced by <math>0.22 \pm 0.07</math> h/day (<math>p = 0.0014</math>) for intervention participants compared with controls; 53% reduction</li> <li>• &lt;45 mg/dL reduced by <math>(-0.14 \pm 0.04)</math> h/day 64% reduction</li> <li>• Nocturnal hypoglycemia 70mg/dL reduced by 54% (<math>-0.29 \pm 0.08</math> h per 7 h) for intervention participants compared with control (<math>p = 0.0001</math>).</li> <li>• Haak T, Hanaire H, Ajjan R et al. Flash Glucose-Sensing Technology as a Replacement for Blood Glucose Monitoring for the Management of Insulin-Treated Type 2 Diabetes: a Multicenter, Open-Label Randomized Controlled Trial. <i>Diabetes Ther</i> (2017 Feb) 8(1):55-73. PMID: 28000140</li> </ul>
2	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes	Use of Flash GM has shown reduction in hypoglycemia incidence. Use of Flash GM in type 2 DM on multi dose insulin regimen is widely covered by Medicare and VA facilities. A provision could be made to cover the Flash/Intermittent GM for type 2 DM on multi dose insulin regimen.
3	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes	Persons with T2DM have highly variable physiology and treatment requirements. Some patients are highly insulin deficient, require physiologic insulin replacement, and are at risk for both severe hyperglycemic and hypoglycemic events. Patients with T2DM who require frequent glucose monitoring to adjust insulin therapy benefit from CGM in a similar manner to persons with T1DM. Persons at higher risk of hypoglycemia may benefit from long-term CGM. The data both less and less compelling for patients who do not require intensified insulin therapy. This group includes those who take basal insulin in addition to non-insulin therapies, and those who take only non-insulin medications for control of their diabetes. While the major problem is hyperglycemia and not hypoglycemia, CGM may be a cost-effective way to improve HbA1c in those not at goal, even reduce medication use. Unfortunately, data for use as an adjunct to lifestyle management is limited. Chehregosha and colleagues discuss the limitations of HbA1c, and greater utility of the metric, glucose management indicator (GMI, previously also known as eA1c) in the management of persons with T2DM. <ul style="list-style-type: none"> <li>• Chehregosha H, Khamseh ME, Malek M, et al. A View Beyond HbA1c: Role of Continuous Glucose Monitoring. <i>Diabetes Ther</i>. 2019 Jun;10(3):853-863. PMID 31037553</li> </ul> Long-term use of CGM should be considered in patients whose diagnostic CGM shows discordance with HbA1c, those at risk for hypoglycemia, those on intensified insulin regimens and those who have achieved proven benefit from short-term CGM. Short term CGM trial for 1-2 weeks is a common practice among endocrinologists in the following situations: <ul style="list-style-type: none"> <li>• In patients who has frequent hypoglycemia episodes</li> <li>• In patients with discrepancy between A1c and Self monitored blood glucose data</li> <li>• Uncontrolled diabetes</li> <li>• To monitor the trends of hyperglycemia and hypoglycemia to guide insulin dosing decisions.</li> <li>• To screen for occult hypoglycemia episodes in patients with hypoglycemic awareness.</li> <li>• Preoperative optimization of glucose control before surgery</li> <li>• Post-operative optimization of glucose control after surgery</li> <li>• During pregnancy and to optimize glucose levels in the preconception period</li> </ul>
1	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes	<ul style="list-style-type: none"> <li>• In patients who has frequent hypoglycemia episodes</li> <li>• In patients with discrepancy between A1c and Self monitored blood glucose data</li> <li>• Uncontrolled diabetes</li> <li>• To monitor the trends of hyperglycemia and hypoglycemia to guide insulin dosing decisions.</li> <li>• To screen for occult hypoglycemia episodes in patients with hypoglycemic awareness.</li> <li>• Preoperative optimization of glucose control before surgery</li> <li>• Post-operative optimization of glucose control after surgery</li> <li>• During pregnancy and to optimize glucose levels in the preconception period</li> </ul>
2	Use of short-term (intermittent) glucose	Based on the available evidence, endocrine society guidelines recommend short-term, intermittent RT-CGM use in adult patients with T2DM (not on prandial insulin) who have A1C levels $\geq 7\%$ and are willing and able to use the device. I agree with this recommendation.

#	Indications	Rationale-CI
	monitoring for individuals with type 2 diabetes	<ul style="list-style-type: none"> <li>Peters AL, Ahmann AJ, Battelino T, et al. Diabetes Technology—Continuous Subcutaneous Insulin Infusion Therapy and Continuous Glucose Monitoring in Adults: An Endocrine Society Clinical Practice Guideline. J Clin Endocrinol Metab. 2016 101(11):3922-3937. PMID: 27588440</li> </ul>
3	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes	Data is very limited for the use of short-term (intermittent) glucose monitoring in persons with T2DM. Devices designed for short-term use are generally used for diagnostic purposes and are blinded to the patient. The downloaded data is used by clinicians to highlight problems and refine treatment. These diagnostic, or professional devices, are available, but are not suited for intermittent use, since either the patient or the device has to be returned to the center for download, and information is generally not available in real time. Current devices are better suited for continuous use, and involve the patient using a phone or dedicated reader to access glucose information. Experience with self-monitoring of blood glucose (SMBG) and with CGM in T1DM suggests that other than diagnostic use, CGM or other monitoring should be an integral part of diabetes care and ongoing.

NR = not reported

- Based on the evidence and your clinical experience for each of the use of **short-term (intermittent)** glucose monitoring for individuals with **type 1 diabetes** with the defined patient selection criteria described below:
  - Respond YES or NO for each clinical indication whether the intervention would be expected to provide a clinically meaningful improvement in net health outcome; AND
  - Rate your level of confidence in your YES or NO response using the 1 to 5 scale outlined below.

#	Indications	YES / NO	Confidence				
			Low Confidence 1	2	Intermediate Confidence 3	4	High Confidence 5
1	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes which is poorly controlled[1], despite current use of best practices and who are willing and able to use the device.	Yes					X
	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes					X
2	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes which is poorly controlled <sup>1</sup> , despite current use of best practices and who are willing and able to use the device.	Yes	X				
	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes			X		
3	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes which is poorly controlled <sup>1</sup> , despite current use of best practices and who are willing and able to use the device.	No			X		
	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes prior to insulin pump initiation to	NR					

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	

determine basal insulin levels and who are willing and able to use the device.

NR = not reported

- Based on the evidence and your clinical experience for each of the use of **short-term (intermittent)** glucose monitoring for individuals with **type 1 diabetes** with the defined patient selection criteria described below:
  - Respond YES or NO for each clinical indication whether this intervention is consistent with generally accepted medical practice; AND
  - Rate your level of confidence in your YES or NO response using the 1 to 5 scale outlined below.

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
1	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes which is poorly controlled <sup>1</sup> , despite current use of best practices and who are willing and able to use the device.	Yes						X
	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes						X
2	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes which is poorly controlled <sup>1</sup> , despite current use of best practices and who are willing and able to use the device.	Yes	X					
	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes		X				
3	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes which is poorly controlled <sup>1</sup> , despite current use of best practices and who are willing and able to use the device.	NR						
	Use of short-term (intermittent) glucose monitoring for individuals with type 1 diabetes prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	NR						

NR = not reported

- Based on the evidence and your clinical experience for each of the use of **long-term (continuous)** glucose monitoring for individuals with **type 2 diabetes** with the defined patient selection criteria described below:

**1.01.20 Continuous Glucose Monitoring**

Page 52 of 65

- Respond YES or NO for each clinical indication whether the intervention would be expected to provide a clinically meaningful improvement in net health outcome; AND
- Rate your level of confidence in your YES or NO response using the 1 to 5 scale outlined below.

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
1	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>[2]</sup> despite use of best practices and who are willing and able to use the device.	Yes					X	
2	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes					X	
3	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes						X

NR = not reported

5. Based on the evidence and your clinical experience for each of the use of **long-term (continuous)** glucose monitoring for individuals with **type 2 diabetes** with the defined patient selection criteria described below:
  - Respond YES or NO for each clinical indication whether this intervention is consistent with generally accepted medical practice; AND
  - Rate your level of confidence in your YES or NO response using the 1 to 5 scale outlined below.

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
1	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes						X
2	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes					X	

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
3	Use of long-term (continuous) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	NR						

NR = not reported

6. Based on the evidence and your clinical experience for each of the use of **short-term (intermittent)** glucose monitoring for individuals with **type 2 diabetes** with the defined patient selection criteria described below:
- Respond YES or NO for each clinical indication whether the intervention would be expected to provide a clinically meaningful improvement in net health outcome; AND
  - Rate your level of confidence in your YES or NO response using the 1 to 5 scale outlined below.

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
1	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes					X	
	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes					X	
2	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes						X
	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes						X
3	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	No			X			

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	NR						

NR = not reported

- Based on the evidence and your clinical experience for each of the use of **short-term (intermittent)** glucose monitoring for individuals with **type 2 diabetes** with the defined patient selection criteria described below:
  - Respond YES or NO for each clinical indication whether this intervention is consistent with generally accepted medical practice; AND
  - Rate your level of confidence in your YES or NO response using the 1 to 5 scale outlined below.

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
1	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes					X	
	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes					X	
2	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.	Yes						X
	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.	Yes						X
3	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin but who are poorly controlled <sup>2</sup> despite use of best practices and who are willing and able to use the device.							

#	Indications	YES / NO	Low Confidence		Intermediate Confidence		High Confidence	
			1	2	3	4	5	
	Use of short-term (intermittent) glucose monitoring for individuals with type 2 diabetes who are insulinopenic and require multiple daily injections of insulin prior to insulin pump initiation to determine basal insulin levels and who are willing and able to use the device.							

NR = not reported

- Additional narrative rationale or comments regarding the clinical context or specific clinical pathways for this topic and/or any relevant scientific citations (including the PMID) with evidence that demonstrates health outcomes you would like to highlight.

#	Additional Comments
1	<ol style="list-style-type: none"> <li>Hypoglycemia is a limiting factor for aggressive insulin dose adjustments, robust real time data offered by CGM will help guide physicians to adjust the insulin regimen accurately and will aid patients to comply with the recommended regimen appropriately, as patients are reassured by the continuous surveillance from CGM.</li> <li>Limitations of using A1c as an outcome measure: <ul style="list-style-type: none"> <li>HbA1c reflects average measurements- and may mask large variations in glycemic control</li> <li>A1c does not address hypoglycemia and which is associated with adverse outcomes.</li> <li>Different glycation potentials cause variation in the A1c levels for the same glucose levels.</li> </ul> </li> </ol> <p>Increasingly it is being recognized A1c by itself is not the end outcome measure and alternative metrics like glucose time in range and glycemic variability should be used. CGM plays an important role in evaluating these metrics</p> <ul style="list-style-type: none"> <li>Sagar R, Abbas A, Ajjan R. Glucose monitoring in diabetes: from clinical studies to real-world practice. Practical Diabetes. March 2019 accessible at URL: <a href="https://www.practicaldiabetes.com/article/glucose-monitoring-in-diabetes-from-clinical-studies-to-real-world-practice/">https://www.practicaldiabetes.com/article/glucose-monitoring-in-diabetes-from-clinical-studies-to-real-world-practice/</a></li> <li>CGMS offers a practical way for diabetes patient to comply with glucose check monitoring recommendations and decreases the disease burden of the patients and improving patient quality of life both by making glucose testing easier and painless</li> <li>Due to the above reasons CGM should be offered as an option for all patients with Type1 and Type2 diabetes patients on intensive insulin treatment. <ul style="list-style-type: none"> <li>Mancini G, Berioli MG, Santi E, et al. Flash Glucose Monitoring: A Review of the Literature with a Special Focus on Type 1 Diabetes. Nutrients. 2018 Aug;10(8):992. PMID: 30060632</li> <li>Sagar R, Abbas A, Ajjan R. Glucose monitoring in diabetes: from clinical studies to real-world practice. Practical Diabetes. March 2019 accessible at URL: <a href="https://www.practicaldiabetes.com/article/glucose-monitoring-in-diabetes-from-clinical-studies-to-real-world-practice/">https://www.practicaldiabetes.com/article/glucose-monitoring-in-diabetes-from-clinical-studies-to-real-world-practice/</a></li> </ul> </li> </ul>
2	Not applicable
3	Two facts are indisputable: first that hyperglycemia over time causes microvascular complications that can shorten the quantity and quality of life; and second that persons who take exogenous insulin are at risk for hypoglycemic events. Glucose monitoring plays an important role in reducing both of these risks. Continuous glucose monitoring has been shown to increase "time in range" (generally glucoses of 70 - 180 mg/dl) and reduce time spent in hypoglycemia. There is no question that persons with T1DM should have access to CGM, despite their current HbA1c. Patients with T2DM who are insulin-treated should also have access to continuous CGM for safety and to help achieve glycemic goals. Given the lower glucose targets in pregnancy, any woman who requires insulin during pregnancy, whether T1DM, T2DM or gestational diabetes, should also have access to CGM. Patients with T2DM who are not at risk for hypoglycemia and who currently do SMBG infrequently may find that the increased information provided by continuous use of CGM helps with lifestyle choices and

#	Additional Comments
	modifications in a cost-effective way, however it is not needed for safety. Children and others who require insulin or who are otherwise at risk for hypoglycemia, and who have a responsible caretaker, should have access to CGM with a data-share feature so that the responsible caretaker can monitor glucose levels. Current CGM systems are not substantially more difficult to use than SMBG, however diabetes education or instruction in device use should be available to those not familiar with the technology.

NR = not reported

9. Is there any evidence missing from the attached draft review of evidence that demonstrates clinically meaningful improvement in net health outcome? If YES, please share any relevant scientific citations of missing evidence (including the PMID).

#	YES / NO	Citations of Missing Evidence
1	Yes	<ul style="list-style-type: none"> <li>Anjana RM, Kesavadev J, Neeta D, et al. A Multicenter Real-Life Study on the Effect of Flash Glucose Monitoring on Glycemic Control in Patients with Type 1 and Type 2 Diabetes. <i>Diabetes Technol Ther.</i> 2017 Sep;19(9):533-540. PMID: 28930495</li> <li>Landau Z, Abiri S, Gruber N et al. Use of flash glucose-sensing technology (FreeStyle Libre) in youth with type 1 diabetes: AWeSoMe study group real-life observational experience. <i>Acta Diabetol.</i> 2018 Dec;55(12):1303-1310. PMID: 30171412</li> <li>Yaron M, Roitman E, Aharon-Hananel G, et al. Effect of Flash Glucose Monitoring Technology on Glycemic Control and Treatment Satisfaction in Patients With Type 2 Diabetes. <i>Diabetes Care.</i> 2019; pii:dc180166. [Epub ahead of print] PMID 31036546</li> <li>Dunn TC, Xu Y, Hayter G et al. Real-world flash glucose monitoring patterns and associations between self-monitoring frequency and glycaemic measures: A European analysis of over 60 million glucose tests. <i>Diabetes Res Clin Pract.</i> 2018 Mar;137:37-46. PMID 29278709</li> <li>Haak T, Hanaire H, Ajjan R et al. Flash Glucose-Sensing Technology as a Replacement for Blood Glucose Monitoring for the Management of Insulin-Treated Type 2 Diabetes: a Multicenter, Open-Label Randomized Controlled Trial. <i>Diabetes Ther</i> (2017 Feb) 8(1):55-73. PMID: 28000140</li> </ul>
2	No	<p>Not applicable</p> <p>Additional information to consider:</p> <p>CGM has been shown to improve neonatal outcomes in women with T1DM, reducing the incidence of large for gestational age (odds ratio 0.51, 95% CI 0.28 – 0.90, p=0.021); fewer neonatal intensive care admissions lasting more than 24 hours (odds ratio 0.48; 0.26 – 0.86; p=0.00157); and 1-day shorter length of hospital stay (p=0.0091).</p> <ul style="list-style-type: none"> <li>Feig DS, Donovan LE, Corcoy R, et al. Continuous glucose monitoring in pregnant women with type 1 diabetes (CONCEPTT): a multicentre international randomised controlled trial. <i>Lancet.</i> 2017;390(10110):2347-2359. PMID: 28923465</li> </ul> <p>Continuous CGM use has been shown to reduce health care utilization in persons with T1DM.</p> <ul style="list-style-type: none"> <li>Parkin CG, Graham C, Smolskis J. Continuous Glucose Monitoring Use in Type 1 Diabetes: Longitudinal Analysis Demonstrates Meaningful Improvements in HbA1c and Reductions in Health Care Utilization. <i>J Diabetes Sci Technol.</i> 2017;11(3):522-528. PMID: 28745091</li> </ul> <p>A meta-analysis of real-time and retrospective (intermittent) CGM use showed reduced HbA1c in persons with T2DM.</p> <ul style="list-style-type: none"> <li>Ida S, Kaneko R, Murata K. Utility of Real-Time and Retrospective Continuous Glucose Monitoring in Patients with Type 2 Diabetes Mellitus: A Meta-Analysis of Randomized Controlled Trials. <i>J Diabetes Res.</i> 2019 Jan;15;2019:4684815. PMID: 30775385</li> </ul> <p>Baseline glycated hemoglobin values predict the magnitude of improvement in patients with both T1DM and T2DM, with greater A1c drops in persons with higher A1c.</p> <ul style="list-style-type: none"> <li>Billings LK, Parkin CG, Price D. Baseline Glycated Hemoglobin Values Predict the Magnitude of Glycemic Improvement in Patients with Type 1 and Type 2 Diabetes: Subgroup Analyses from the DIAMOND Study Program. <i>Diabetes Technol Ther.</i> 2018;20(8):561-565. PMID: 30044123</li> </ul>
3	Yes	

[1] Type 1 diabetes which is poorly controlled includes the following clinical situations: unexplained hypoglycemic episodes, hypoglycemic unawareness, suspected postprandial hyperglycemia, and

recurrent diabetic ketoacidosis.

[2] Type 2 diabetes which is poorly controlled includes the following clinical situations: unexplained hypoglycemic episodes, hypoglycemic unawareness, and persistent hyperglycemia and A1C levels above target.

## References

1. Writing Team for the Diabetes Control and Complications Trial/Epidemiology of Diabetes Interventions and Complications Research Group. Effect of intensive therapy on the microvascular complications of type 1 diabetes mellitus. *JAMA*. May 15 2002; 287(19): 2563-9. PMID 12020338
2. American Diabetes Association. 13. Management of Diabetes in Pregnancy: Standards of Medical Care in Diabetes-2018. *Diabetes Care*. Jan 2018; 41(Suppl 1): S137-S143. PMID 29222384
3. Pazos-Couselo M, Garcia-Lopez JM, Gonzalez-Rodriguez M, et al. High incidence of hypoglycemia in stable insulin-treated type 2 diabetes mellitus: continuous glucose monitoring vs. self-monitored blood glucose. Observational prospective study. *Can J Diabetes*. Oct 2015; 39(5): 428-33. PMID 26254702
4. Gehlaut RR, Dogbey GY, Schwartz FL, et al. Hypoglycemia in Type 2 Diabetes--More Common Than You Think: A Continuous Glucose Monitoring Study. *J Diabetes Sci Technol*. Apr 27 2015; 9(5): 999-1005. PMID 25917335
5. Food and Drug Administration (FDA). Summary of Safety and Effectiveness (SSED): Dexcom G5 Mobile Continuous Glucose Monitoring System. 2016; <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpma/pma.cfm?id=P120005S041>. Accessed November 27, 2021.
6. Floyd B, Chandra P, Hall S, et al. Comparative analysis of the efficacy of continuous glucose monitoring and self-monitoring of blood glucose in type 1 diabetes mellitus. *J Diabetes Sci Technol*. Sep 01 2012; 6(5): 1094-102. PMID 23063035
7. Gandhi GY, Kovalaske M, Kudva Y, et al. Efficacy of continuous glucose monitoring in improving glycemic control and reducing hypoglycemia: a systematic review and meta-analysis of randomized trials. *J Diabetes Sci Technol*. Jul 01 2011; 5(4): 952-65. PMID 21880239
8. Langendam M, Luyf YM, Hooft L, et al. Continuous glucose monitoring systems for type 1 diabetes mellitus. *Cochrane Database Syst Rev*. Jan 18 2012; 1: CD008101. PMID 22258980
9. Poolsup N, Suksomboon N, Kyaw AM. Systematic review and meta-analysis of the effectiveness of continuous glucose monitoring (CGM) on glucose control in diabetes. *Diabetol Metab Syndr*. 2013; 5: 39. PMID 23876067
10. Wojciechowski P, Rys P, Lipowska A, et al. Efficacy and safety comparison of continuous glucose monitoring and self-monitoring of blood glucose in type 1 diabetes: systematic review and meta-analysis. *Pol Arch Med Wewn*. Oct 2011; 121(10): 333-43. PMID 22045094
11. Yeoh E, Choudhary P, Nwokolo M, et al. Interventions That Restore Awareness of Hypoglycemia in Adults With Type 1 Diabetes: A Systematic Review and Meta-analysis. *Diabetes Care*. Aug 2015; 38(8): 1592-609. PMID 26207053
12. Benkhadra K, Alahdab F, Tamhane S, et al. Real-time continuous glucose monitoring in type 1 diabetes: a systematic review and individual patient data meta-analysis. *Clin Endocrinol (Oxf)*. Mar 2017; 86(3): 354-360. PMID 27978595
13. Lind M, Polonsky W, Hirsch IB, et al. Continuous Glucose Monitoring vs Conventional Therapy for Glycemic Control in Adults With Type 1 Diabetes Treated With Multiple Daily Insulin Injections: The GOLD Randomized Clinical Trial. *JAMA*. Jan 24 2017; 317(4): 379-387. PMID 28118454
14. Beck RW, Riddlesworth T, Ruedy K, et al. Effect of Continuous Glucose Monitoring on Glycemic Control in Adults With Type 1 Diabetes Using Insulin Injections: The DIAMOND Randomized Clinical Trial. *JAMA*. Jan 24 2017; 317(4): 371-378. PMID 28118453

15. Riddlesworth T, Price D, Cohen N, et al. Hypoglycemic Event Frequency and the Effect of Continuous Glucose Monitoring in Adults with Type 1 Diabetes Using Multiple Daily Insulin Injections. *Diabetes Ther.* Aug 2017; 8(4): 947-951. PMID 28616804
16. Polonsky WH, Hessler D, Ruedy KJ, et al. The Impact of Continuous Glucose Monitoring on Markers of Quality of Life in Adults With Type 1 Diabetes: Further Findings From the DIAMOND Randomized Clinical Trial. *Diabetes Care.* Jun 2017; 40(6): 736-741. PMID 28389582
17. Laffel LM, Kanapka LG, Beck RW, et al. Effect of Continuous Glucose Monitoring on Glycemic Control in Adolescents and Young Adults With Type 1 Diabetes: A Randomized Clinical Trial. *JAMA.* Jun 16 2020; 323(23): 2388-2396. PMID 32543683
18. Pratley RE, Kanapka LG, Rickels MR, et al. Effect of Continuous Glucose Monitoring on Hypoglycemia in Older Adults With Type 1 Diabetes: A Randomized Clinical Trial. *JAMA.* Jun 16 2020; 323(23): 2397-2406. PMID 32543682
19. Feig DS, Donovan LE, Corcoy R, et al. Continuous glucose monitoring in pregnant women with type 1 diabetes (CONCEPTT): a multicentre international randomised controlled trial. *Lancet.* Nov 25 2017; 390(10110): 2347-2359. PMID 28923465
20. Kropff J, Choudhary P, Neupane S, et al. Accuracy and Longevity of an Implantable Continuous Glucose Sensor in the PRECISE Study: A 180-Day, Prospective, Multicenter, Pivotal Trial. *Diabetes Care.* Jan 2017; 40(1): 63-68. PMID 27815290
21. Christiansen MP, Klaff LJ, Brazg R, et al. A Prospective Multicenter Evaluation of the Accuracy of a Novel Implanted Continuous Glucose Sensor: PRECISE II. *Diabetes Technol Ther.* Mar 2018; 20(3): 197-206. PMID 29381090
22. Food and Drug Administration (FDA). Summary of Safety and Effectiveness Data: Eversense Continuous Glucose Monitoring System (2018). <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpma/pma.cfm?id=P160048>. Accessed October November 26, 2021.
23. Sanchez P, Ghosh-Dastidar S, Tweden KS, et al. Real-World Data from the First U.S. Commercial Users of an Implantable Continuous Glucose Sensor. *Diabetes Technol Ther.* Dec 2019; 21(12): 677-681. PMID 31385732
24. Deiss D, Irace C, Carlson G, et al. Real-World Safety of an Implantable Continuous Glucose Sensor Over Multiple Cycles of Use: A Post-Market Registry Study. *Diabetes Technol Ther.* Jan 2020; 22(1): 48-52. PMID 31418587
25. Tweden KS, Deiss D, Rastogi R, et al. Longitudinal Analysis of Real-World Performance of an Implantable Continuous Glucose Sensor over Multiple Sensor Insertion and Removal Cycles. *Diabetes Technol Ther.* May 2020; 22(5): 422-427. PMID 31697182
26. Newman SP, Cooke D, Casbard A, et al. A randomised controlled trial to compare minimally invasive glucose monitoring devices with conventional monitoring in the management of insulin-treated diabetes mellitus (MITRE). *Health Technol Assess.* May 2009; 13(28): iii-iv, ix-xi, 1-194. PMID 19476724
27. Voormolen DN, DeVries JH, Evers IM, et al. The efficacy and effectiveness of continuous glucose monitoring during pregnancy: a systematic review. *Obstet Gynecol Surv.* Nov 2013; 68(11): 753-63. PMID 24193194
28. Secher AL, Ringholm L, Andersen HU, et al. The effect of real-time continuous glucose monitoring in pregnant women with diabetes: a randomized controlled trial. *Diabetes Care.* Jul 2013; 36(7): 1877-83. PMID 23349548
29. Murphy HR, Rayman G, Lewis K, et al. Effectiveness of continuous glucose monitoring in pregnant women with diabetes: randomised clinical trial. *BMJ.* Sep 25 2008; 337: a1680. PMID 18818254
30. Ida S, Kaneko R, Murata K. Utility of Real-Time and Retrospective Continuous Glucose Monitoring in Patients with Type 2 Diabetes Mellitus: A Meta-Analysis of Randomized Controlled Trials. *J Diabetes Res.* 2019; 2019: 4684815. PMID 30775385
31. Ehrhardt NM, Chellappa M, Walker MS, et al. The effect of real-time continuous glucose monitoring on glycemic control in patients with type 2 diabetes mellitus. *J Diabetes Sci Technol.* May 01 2011; 5(3): 668-75. PMID 21722581
32. Cosson E, Hamo-Tchatchouang E, Dufaitre-Patouraux L, et al. Multicentre, randomised, controlled study of the impact of continuous sub-cutaneous glucose monitoring

- (GlucoDay) on glycaemic control in type 1 and type 2 diabetes patients. *Diabetes Metab. Sep 2009*; 35(4): 312-8. PMID 19560388
33. Allen NA, Fain JA, Braun B, et al. Continuous glucose monitoring counseling improves physical activity behaviors of individuals with type 2 diabetes: A randomized clinical trial. *Diabetes Res Clin Pract. Jun 2008*; 80(3): 371-9. PMID 18304674
  34. Yoo HJ, An HG, Park SY, et al. Use of a real time continuous glucose monitoring system as a motivational device for poorly controlled type 2 diabetes. *Diabetes Res Clin Pract. Oct 2008*; 82(1): 73-9. PMID 18701183
  35. Ajjan RA, Abougila K, Bellary S, et al. Sensor and software use for the glycaemic management of insulin-treated type 1 and type 2 diabetes patients. *Diab Vasc Dis Res. May 2016*; 13(3): 211-9. PMID 27000105
  36. Haak T, Hanaire H, Ajjan R, et al. Flash Glucose-Sensing Technology as a Replacement for Blood Glucose Monitoring for the Management of Insulin-Treated Type 2 Diabetes: a Multicenter, Open-Label Randomized Controlled Trial. *Diabetes Ther. Feb 2017*; 8(1): 55-73. PMID 28000140
  37. Beck RW, Riddlesworth TD, Ruedy K, et al. Continuous Glucose Monitoring Versus Usual Care in Patients With Type 2 Diabetes Receiving Multiple Daily Insulin Injections: A Randomized Trial. *Ann Intern Med. Sep 19 2017*; 167(6): 365-374. PMID 28828487
  38. Vigersky RA, Fonda SJ, Chellappa M, et al. Short- and long-term effects of real-time continuous glucose monitoring in patients with type 2 diabetes. *Diabetes Care. Jan 2012*; 35(1): 32-8. PMID 22100963
  39. Martens T, Beck RW, Bailey R, et al. Effect of Continuous Glucose Monitoring on Glycemic Control in Patients With Type 2 Diabetes Treated With Basal Insulin: A Randomized Clinical Trial. *JAMA. Jun 08 2021*; 325(22): 2262-2272. PMID 34077499
  40. Haak T, Hanaire H, Ajjan R, et al. Use of Flash Glucose-Sensing Technology for 12 months as a Replacement for Blood Glucose Monitoring in Insulin-treated Type 2 Diabetes. *Diabetes Ther. Jun 2017*; 8(3): 573-586. PMID 28401454
  41. Wei Q, Sun Z, Yang Y, et al. Effect of a CGMS and SMBG on Maternal and Neonatal Outcomes in Gestational Diabetes Mellitus: a Randomized Controlled Trial. *Sci Rep. Jan 27 2016*; 6: 19920. PMID 26814139
  42. Bolinder J, Antuna R, Geelhoed-Duijvestijn P, et al. Novel glucose-sensing technology and hypoglycaemia in type 1 diabetes: a multicentre, non-masked, randomised controlled trial. *Lancet. Nov 05 2016*; 388(10057): 2254-2263. PMID 27634581
  43. Secher AL, Pedersen-Bjergaard U, Svendsen OL, et al. Flash glucose monitoring and automated bolus calculation in type 1 diabetes treated with multiple daily insulin injections: a 26 week randomised, controlled, multicentre trial. *Diabetologia. Dec 2021*; 64(12): 2713-2724. PMID 34495375
  44. Yaron M, Roitman E, Aharon-Hananel G, et al. Effect of Flash Glucose Monitoring Technology on Glycemic Control and Treatment Satisfaction in Patients With Type 2 Diabetes. *Diabetes Care. Jul 2019*; 42(7): 1178-1184. PMID 31036546
  45. Furler J, O'Neal D, Speight J, et al. Use of professional-mode flash glucose monitoring, at 3-month intervals, in adults with type 2 diabetes in general practice (GP-OSMOTIC): a pragmatic, open-label, 12-month, randomised controlled trial. *Lancet Diabetes Endocrinol. Jan 2020*; 8(1): 17-26. PMID 31862147
  46. American Association of Clinical Endocrinology and American College of Endocrinology. Comprehensive Type 2 Diabetes Management Algorithm. 2020. <https://pro.aace.com/disease-state-resources/diabetes/clinical-practice-guidelines-treatment-algorithms/comprehensive>. Accessed November 29, 2021.
  47. National Institute for Health and Care Excellence (NICE). Type 1 diabetes in adults: diagnosis and management [NG17]. 2016; <https://www.nice.org.uk/guidance/ng17?unlid=382286372016220232952>. Accessed November 29, 2021.
  48. American Diabetes Association. Standards of Medical Care in Diabetes. 2021. <https://professional.diabetes.org/content-page/practice-guidelines-resources>. Accessed November 29, 2021.

49. Peters AL, Ahmann AJ, Battelino T, et al. Diabetes Technology-Continuous Subcutaneous Insulin Infusion Therapy and Continuous Glucose Monitoring in Adults: An Endocrine Society Clinical Practice Guideline. *J Clin Endocrinol Metab.* Nov 2016; 101(11): 3922-3937. PMID 27588440
50. Battelino T, Danne T, Bergenstal RM, et al. Clinical Targets for Continuous Glucose Monitoring Data Interpretation: Recommendations From the International Consensus on Time in Range. *Diabetes Care.* Aug 2019; 42(8): 1593-1603. PMID 31177185
51. Centers for Medicare & Medicare Services. Durable Medical Equipment, Prosthetics/Orthotics & Supplies Fee Schedule. <https://www.cms.gov/medicare/medicare-fee-for-service-payment/dmeposfeesched>. Accessed November 28, 2021.
52. Centers for Medicare & Medicare Services. Durable Medical Equipment (DME) Center; <https://www.cms.gov/Center/Provider-Type/Durable-Medical-Equipment-DME-Center>. Accessed November 29, 2021.

## Documentation for Clinical Review

### Please provide the following documentation:

#### Initial Request:

- History and physical and/or consultation notes from referring physician including:
  - Type of diabetes and duration, reason for the request
  - Provider attestation that the patient has insulin dependent (type 1 or type 2) diabetes requiring multiple daily doses of insulin; that they have demonstrated an understanding of the technology, are motivated to use the device correctly and consistently, are expected to adhere to a comprehensive diabetes treatment plan supervised by a qualified provider, and are capable of using the device to recognize alerts and alarms
  - Current insulin therapy and recent adjustments
  - Reason for short term need if appropriate
- Documented frequency of glucose self-testing and number of insulin injections per day or self-adjustments on an insulin pump (i.e., blood sugar and insulin logs), for the past 30 days to support the provider attestation
- Type (name) of device being requested

#### Replacements and/or Repair:

- Clinical summary including:
  - Type of diabetes and insulin management
  - Past benefit from CGM device, including clinical findings
  - Reason for continued need of CGM device
  - Description of device malfunction
- Warranty information and repair log or repair history (if applicable)

### Post Service (in addition to the above, please include the following):

- All requirements for an initial request, plus:
  - Results/reports of blood sugar and insulin logs performed or device report of data

## Coding

*This Policy relates only to the services or supplies described herein. Benefits may vary according to product design; therefore, contract language should be reviewed before applying the terms of the Policy.*

*The following codes are included below for informational purposes. Inclusion or exclusion of a code(s) does not constitute or imply member coverage or provider reimbursement policy. Policy Statements are intended to provide member coverage information and may*

include the use of some codes for clarity. The Policy Guidelines section may also provide additional information for how to interpret the Policy Statements and to provide coding guidance in some cases.

Type	Code	Description
CPT®	95249	Ambulatory continuous glucose monitoring of interstitial tissue fluid via a subcutaneous sensor for a minimum of 72 hours; patient-provided equipment, sensor placement, hook-up, calibration of monitor, patient training, and printout of recording
	95250	Ambulatory continuous glucose monitoring of interstitial tissue fluid via a subcutaneous sensor for a minimum of 72 hours; physician or other qualified health care professional (office) provided equipment, sensor placement, hook-up, calibration of monitor, patient training, removal of sensor, and printout of recording
	95251	Ambulatory continuous glucose monitoring of interstitial tissue fluid via a subcutaneous sensor for a minimum of 72 hours; analysis, interpretation and report
	99091	Collection and interpretation of physiologic data (e.g., ECG, blood pressure, glucose monitoring) digitally stored and/or transmitted by the patient and/or caregiver to the physician or other qualified health care professional, qualified by education, training, licensure/regulation (when applicable) requiring a minimum of 30 minutes of time, each 30 days
	0446T	Creation of subcutaneous pocket with insertion of implantable interstitial glucose sensor, including system activation and patient training
	0447T	Removal of implantable interstitial glucose sensor from subcutaneous pocket via incision
	0448T	Removal of implantable interstitial glucose sensor with creation of subcutaneous pocket at different anatomic site and insertion of new implantable sensor, including system activation
HCPCS	A4226	Supplies for maintenance of insulin infusion pump with dosage rate adjustment using therapeutic continuous glucose sensing, per week
	A4238	Supply allowance for adjunctive continuous glucose monitor (CGM), includes all supplies and accessories, 1 month supply = 1 unit of service <b>(Code effective 4/1/2022)</b>
	A9276	Sensor; invasive (e.g., subcutaneous), disposable, for use with interstitial continuous glucose monitoring system, 1 unit = 1 day supply
	A9277	Transmitter; external, for use with interstitial continuous glucose monitoring system
	A9278	Receiver (monitor); external, for use with interstitial continuous glucose monitoring system
	E0787	External ambulatory infusion pump, insulin, dosage rate adjustment using therapeutic continuous glucose sensing
	E2102	Adjunctive continuous glucose monitor or receiver <b>(Code effective 4/1/2022)</b>
	K0553	Supply allowance for therapeutic continuous glucose monitor (CGM), includes all supplies and accessories, 1 month supply = 1 unit of service
	K0554	Receiver (monitor), dedicated, for use with therapeutic glucose continuous monitor system
S1030	Continuous noninvasive glucose monitoring device, purchase (for physician interpretation of data, use CPT code)	

Type	Code	Description
	S1031	Continuous noninvasive glucose monitoring device, rental, including sensor, sensor replacement, and download to monitor (for physician interpretation of data, use CPT code)

### Policy History

This section provides a chronological history of the activities, updates and changes that have occurred with this Medical Policy.

Effective Date	Action
03/01/2020	New policy. Coding Update.
02/01/2021	Annual review. Policy statement, guidelines and literature updated.
02/01/2022	Annual review. Policy statement, guidelines and literature updated. Policy title changed from Continuous or Intermittent Monitoring of Glucose in the Interstitial Fluid to current one.
06/01/2022	Annual review. Policy statement and guidelines updated. Coding Update.

### Definitions of Decision Determinations

**Medically Necessary:** Services that are Medically Necessary include only those which have been established as safe and effective, are furnished under generally accepted professional standards to treat illness, injury or medical condition, and which, as determined by Blue Shield, are: (a) consistent with Blue Shield medical policy; (b) consistent with the symptoms or diagnosis; (c) not furnished primarily for the convenience of the patient, the attending Physician or other provider; (d) furnished at the most appropriate level which can be provided safely and effectively to the patient; and (e) not more costly than an alternative service or sequence of services at least as likely to produce equivalent therapeutic or diagnostic results as to the diagnosis or treatment of the Member's illness, injury, or disease.

**Investigational/Experimental:** A treatment, procedure, or drug is investigational when it has not been recognized as safe and effective for use in treating the particular condition in accordance with generally accepted professional medical standards. This includes services where approval by the federal or state governmental is required prior to use, but has not yet been granted.

**Split Evaluation:** Blue Shield of California/Blue Shield of California Life & Health Insurance Company (Blue Shield) policy review can result in a split evaluation, where a treatment, procedure, or drug will be considered to be investigational for certain indications or conditions, but will be deemed safe and effective for other indications or conditions, and therefore potentially medically necessary in those instances.

### Prior Authorization Requirements (as applicable to your plan)

Within five days before the actual date of service, the provider must confirm with Blue Shield that the member's health plan coverage is still in effect. Blue Shield reserves the right to revoke an authorization prior to services being rendered based on cancellation of the member's eligibility. Final determination of benefits will be made after review of the claim for limitations or exclusions.

Questions regarding the applicability of this policy should be directed to the Prior Authorization Department at (800) 541-6652, or the Transplant Case Management Department at (800) 637-2066 ext. 3507708 or visit the provider portal at [www.blueshieldca.com/provider](http://www.blueshieldca.com/provider).

*Disclaimer: This medical policy is a guide in evaluating the medical necessity of a particular service or treatment. Blue Shield of California may consider published peer-reviewed scientific literature, national*

## **1.01.20 Continuous Glucose Monitoring**

Page 63 of 65

*guidelines, and local standards of practice in developing its medical policy. Federal and state law, as well as contract language, including definitions and specific contract provisions/exclusions, take precedence over medical policy and must be considered first in determining covered services. Member contracts may differ in their benefits. Blue Shield reserves the right to review and update policies as appropriate.*

**Appendix A**

POLICY STATEMENT	
BEFORE <b>Red font: Verbiage removed</b>	AFTER <b>Blue font: Verbiage Changes/Additions</b>
<p><b>Continuous Glucose Monitoring 1.01.20</b></p> <p><b>Policy Statement:</b>                      Short-term (usually less than 72 hours) or long-term <u>real time</u> (true continuous) Glucose Monitoring (rtCGM) devices (e.g., Dexcom) of glucose levels in interstitial fluid, as a technique of diabetic monitoring, may be considered <b>medically necessary</b> when <b>all</b> of the following situations occur:</p> <ol style="list-style-type: none"> <li>I. Patients with insulin dependent (type 1 or type 2) diabetes <b>requiring</b> multiple daily doses of insulin, <b>whose physician attests they have demonstrated an understanding of the technology, are motivated to use the device correctly and consistently, are expected to adhere to a comprehensive diabetes treatment plan supervised by a qualified provider, and are capable of using the device to recognize alerts and alarms</b></li> <li>II. The device includes an audible or tactile (vibrating) alarm for low glucose alerts without patient intervention</li> </ol> <p>Short-term rtCGM monitoring of glucose levels in interstitial fluid may be considered <b>medically necessary</b> in patients with type 1 diabetes mellitus prior to insulin pump initiation to determine basal insulin levels.</p> <p>Repeat short-term CGM may be considered <b>medically necessary</b> in patients who met initial criteria for approval when documentation is submitted to support the need, and including but not limited to why it is preferable to long term CGM or the use of an insulin pump with integrated CGM.</p> <p>The use of <u>intermittently scanned</u> (flash CGM or isCGM) devices (e.g., FreeStyle Libre) are considered <b>investigational</b> (not true continuous monitoring).</p>	<p><b>Continuous Glucose Monitoring 1.01.20</b></p> <p><b>Policy Statement:</b>  <u>Continuous glucose monitoring (CGM)</u> of glucose levels in interstitial fluid, as a technique of diabetic monitoring, may be considered <b>medically necessary</b> when <b>both</b> of the following situations occur:</p> <ol style="list-style-type: none"> <li>I. Patients with insulin dependent (type 1 or type 2) diabetes <b>requires</b> multiple (<b>three or more</b>) daily doses of insulin</li> <li>II. The device includes an audible or tactile (vibrating) alarm for low glucose alerts without patient intervention                      (NOTE: the FreeStyle Libre 14 day device does not have alarms but the FreeStyle Libre 2 does have appropriate alarms, as do Dexcom G5 and G6)</li> </ol>

**POLICY STATEMENT**

<p style="text-align: center;"><b>BEFORE</b></p> <p style="text-align: center;"><b>Red font: Verbiage removed</b></p>	<p style="text-align: center;"><b>AFTER</b></p> <p style="text-align: center;"><b>Blue font: Verbiage Changes/Additions</b></p>
<p>The use of <a href="#">implantable CGM</a> devices (e.g., Eversense®) are considered <b>investigational</b> (see Policy Guidelines section).</p> <p>The use of <a href="#">continuous noninvasive glucose monitoring devices</a> (see Policy Guidelines section) are considered <b>investigational</b>.</p> <p>Other uses of long-term <b>and short-term</b> CGM of glucose levels in interstitial fluid as a technique of diabetic monitoring in patients who are not insulin dependent (including use in gestational diabetes) are considered <b>investigational</b>.</p>	<p>The use of <a href="#">implantable CGM</a> devices (e.g., Eversense®) are considered <b>investigational</b> (see Policy Guidelines section).</p> <p>The use of <a href="#">continuous noninvasive glucose monitoring devices</a> (see Policy Guidelines section) are considered <b>investigational</b>.</p> <p>The use of stand-alone CGMs (e.g., Guardian Connect, that is not part of an insulin infusion pump) <a href="#">adjunctive or non-therapeutic devices</a> are considered <b>investigational</b>.</p> <p>Other uses of long-term CGM of glucose levels in interstitial fluid as a technique of diabetic monitoring in patients who are not insulin dependent (including use in gestational diabetes) are considered <b>investigational</b>.</p>